



inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

GLOBAL TENDER ENQUIRY

2 PART TENDER FOR Light Sheet Microscope-Qty 01 No.

Tender Fee –Nil, EMD: Nil
Cost of Tender: Approx. Rs. 289 Lakhs
Last date for Sale of Documents: 07-02-2022 till 13.00hrs
Last date for submission: 07-02-2022 till 15.00hrs
Due date for opening bids: 07-02-2022 at 15.30hrs

GENERAL CONDITIONS

(Tender documents consist of 27 pages, Pg. 1 to Pg.26 – Technical Bid, Pg. 27- Price Bid)

Important Instructions: The bids shall be enclosed in an envelope and sealed duly marked “Light Sheet Microscope-Qty 01 No.” Ref. No. INS/F-5537/2021-2022 (Y)”; addressed and to be mailed to “The Purchase Officer”. The bids are liable to be rejected if the sealed envelope is not addressed to “The Purchase Officer” with Tender Ref. No. and Item Description. Offers delivered in person shall be deposited in the Tender Box Labelled as ‘TENDER BOX FOR PURCHASE TENDER’ kept in the Ground Floor, Reception at Administration Building. If the bids are sent through courier or mail, it should reach by submission Date and Time and inStem will not responsible for the delay.

Important instruction for the Tenderers: -

All tenders shall be made in ENGLISH only.

The details in regard to technical specification and other terms and conditions should be cogent and clear to the extent possible.

Sealed tenders are invited at Institute for Stem Cell Science and Regenerative Medicine -inStem, National Centre for Biological Sciences, GKVK Post, Bellary Road, Bangalore – 560 065 under 2 cover system from reputed OEM (Original Equipment Manufacturer) or reputed Manufacturers / Authorised Dealers for Supply, Installation, Commissioning and Technical Support for the following: -

Light Sheet Microscope-Qty 01 No.

The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.

D) The EMD amount should be put in **first sealed cover – Cover I**, and super scribed as “**Techno-commercial Bid**” and should contain -

1. Acceptance of Technical specifications, **Annexure- A (page 12)** and terms and conditions. Tender document to be enclosed with Date, Signature and Seal in every Page.
2. Complete Technical details of the Instrument offered (Specifications, Technical Parameters, Advantages, etc.,)
3. Supplier profile & Schedule of Experience – **Annexure – B**
4. Supplier must describe in detail the technical support they will be able to provide in Bangalore. Only those companies will be considered who have engineers based in Bangalore who have been trained on the machines being quoted for, prior to the date of installation.
5. Suppliers must provide complete list of publications arising from use of their machines, in which the machines have been used for the applications listed. Enclose Data Sheet and Sample Analysis (if any)
6. Details of Warranty Services.
7. Details of “After Sales Service” and “Factory Trained Engineers” available in Bangalore.
8. Xerox copies of the Purchase Orders for having supplied similar Instruments in India.
9. Schedule of deviation from specifications / conditions – **Annexure C.**





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

10. Any other information the bidder like to provide with Date, Signature and Seal
(Annexures should be duly signed and filled with date wherever necessary)
(Please attach additional sheet(s), wherever necessary)

II) The **second sealed cover** – Cover II super scribed '**Price Bid**' (page 27) should contain only rates
(please attach additional sheet(s), wherever necessary) (should be duly signed with seal and filled with date wherever necessary)

THE ABOVE MENTIONED TWO COVERS SHALL BE SEALED ON THE OUTSIDE WITH WAX SEAL BEARING THE LOGO/NAME OF THE COMPANY SUBMITTING THE BID.

THESE TWO COVERS SHALL BE AGAIN PUT INTO A SINGLE WAX SEALED COVER super scribed "**Light Sheet Microscope-Qty 01 No.**" and should reach INSTEM on or before **07-02-2022 before 15.00 hrs**". This should be addressed to the Purchase Officer, inStem, NCBS, GKVK Post, Bellary Road, Bangalore – 65. The Techno – Commercial Bid will be opened on **07-02-2022 at 15.30 hrs**. On the date of tender opening (i.e. on **07-02-2022**), only the **Techno-Commercial Bids** shall be opened in the presence of attending tenderers.

The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% and above will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.

MANNER & METHOD FOR SUBMISSION OF BIDS:

1.1 Quotations must be submitted giving complete details using enclosed tender papers.

1.2 The rates quoted should remain valid for a period of 180 days from the date of Price Bid opening.

1.3 Each page of the tender except the Price & Delivery part shall be on printed letterheads or forms and bear the signature, date, name and designation of the person signing the offer. If they are not on letterheads, a rubber stamp indicating full name, address and phone No., Telex No., Fax No. etc. of the firm shall be affixed at the end of each page. The price & delivery part shall be as Annexure B attached.

1.4 This tender document is not transferable. Only the party to whom the tender documents have been issued shall be entitled to quote.

1.5 Bids containing erasures or alterations are liable to be rejected unless countersigned by the Authorised signatory.

1.6 All rates and total amount should be written both in figures and in words and if there is any discrepancy between the two, the lowest amount only will be considered.

1.7 Tenders will be opened in the presence of attending tenderers on **07-02-2022 at 15.30 hours**. In the event the due date is declared as holiday, the tender will be opened on the following working day.





- 1.8 If the item offered is to be imported, arrangements for import will be made by the Institute. Import duty and customs clearance will be under purchaser's scope.
- 1.9 We reserve the right to place order for part/reduced quantity than what is specified in the tender and also reserve the right to split the order to more than one supplier.
- 1.10 Any deviation/substitution in regard to the technical specification must be indicated in Annexure C of this tender document. Otherwise it shall be binding on the bidder to supply the items as specified in this tender specification.
- 1.11 All bids are to be submitted before the due date and time. **The bids shall be enclosed in an envelope and sealed duly marked "Light Sheet Microscope-Qty 01 No.;" Ref. No. INS/F-5537/2021-2022 (Y)"; and mailed to the Purchase Officer.** Offers delivered in person shall be deposited in the Tender Box kept in the office. **Due date for opening the bids: 07-02-2022 (15.00 hours).**
- 1.12 Please return the tender papers including Conditions of Tender as well as the Annexures with your signature, rubber stamp and date affixed on each page.
- 1.13 All bids in response to this invitation of tender should be submitted in a manner and method specified above. Tender which do not comply with the above conditions are liable to be rejected.
- 1.14 Late and delayed tenders will not be considered. Therefore, tenderers shall ensure that the tender reaches the Purchaser on or before the due date and time stipulated for receipt of bids.
TENDERS RECEIVED LATE OR AFTER THE DUE DATE WILL NOT BE CONSIDERED. INSTEM RESERVES THE RIGHT TO ACCEPT, REJECT ANY OR ALL TENDERS WITHOUT ASSIGNING ANY REASONS THERE OF.
- 1.15 Individuals signing the bid form and other supporting documents must specify the capacity in which they sign, like -
- Whether signing as a Sole Proprietor of the firm or his attorney.
 - Whether signing as a partner of the firm or his attorney
 - Whether signing for the firm as Agent.
 - Whether signing as Director of a Limited Company.
- 1.16 The details in regard to technical specification and other terms and conditions should be cogent and clear to the extent possible.
- 1.17 FOLLOWING ANNEXURES must be filled up properly
- | | | |
|--|---|------------|
| (a) Technical Specifications | - | Annexure A |
| (b) Schedule of Experience | - | Annexure B |
| (c) Schedule of deviation from Specifications/conditions | - | Annexure C |
| (d) Schedule of price and delivery | - | Annexure D |
- 1.18 Each page of the tender except the Price & Delivery part shall be on printed letterheads or forms and bear the signature, date, name and designation of the person signing the offer. If they are not on letterheads, a rubber stamp indicating full name, address and phone No., Telex No., Fax No. etc. of the firm shall be affixed at the end of each page. The price & delivery part shall be as Annexure D attached.





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

1.19 All tenders shall be made in ENGLISH only.

2. CATALOGUE/TECHNICAL LITERATURE

All necessary catalogue/drawing literature/data and details of item/s as are considered to be essential for full and correct evaluation of the bid shall invariably accompany the bid.

3. BID GUARANTEE / EARNEST MONEY DEPOSIT: -

Bid Guarantee amount details are as below: -

Sl. No.	Item Description	BGA/EMD Amount
1.	Light Sheet Microscope-Qty 01 No	Nil

4. AUTHORITY OF AGENTS

In case where a bid is submitted by an Indian Agent on behalf of his foreign principal, such bid should be supported with a letter of Authority from the principal that the Indian Agent has been authorised to submit the bid on behalf of the principal. The indigenous manufacturers shall submit their bids directly.

5. PRICE

The tender to be quoted in foreign currencies & any other currencies approved/traded by RBI-USD/Euro/JPY/GBP/SGD/CAD/INR. The price/s quoted shall be firm till the complete execution of the order. All details relating to price, price breakup, inland transportation, documentation, taxes and duties, levies, Road/AIR/Marine freight charges, delivery terms (ex-works/F.O.R/F.O.B/CIP) mode of payment, mode of Dispatch, Insurance, Agency Commission, if any, should be clearly stated. For indicating the price, the tenderers may choose any/all of the following:

- a) Ex-works (all other charges to be indicated separately).
- b) F.O.R. site (i.e. Freight, Packing & Forwarding, loading on to the transport, documentation etc. included.) Internal transportation, Freight, Insurance, etc. to be shown separately.
- c) F.O.B (cost of goods, Packing & Forwarding, Inland Transportation, Documentation, etc. till the item is loaded on to the cargo carrier). Freight & Insurance charges to be separately indicated.
- d) CIP (cost of goods, packing & forwarding, documentation, freight, insurance, etc. all included). However, freight & insurance charges to be indicated separately.

INSURANCE TO BE COVERED TILL inStem STORES.

6. The taxes (State, Central, Turnover tax, Works Contract Tax, etc.). Please specify which are applicable. The duties and other levies, freight, insurance shall be stated clearly and separately. Also please mention whether the same is included in the price/s quoted.

7. VALIDITY OF BIDS

The bids should be valid for a period of 180 days from the date of opening of the Price bids. Bids with shorter validity period are liable for rejection.





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

8. DELIVERY

The tenderer should clearly mention the time required for supplying the item. The period of delivery will be counted (a) from the date of receipt of the order in case of Indian supplier and (b) from the date of opening of letter of credit/receipt of order (in the case of sight draft/advance draft) for foreign suppliers. The delivery date is the date at which the equipment should be delivered at INSTEM.

9. PACKING

The item should be packed appropriately so that it can sustain transit hazards, multiple landing, warehousing, etc. during transit.

10. PLACE OF DELIVERY

In the case of items indigenously manufactured, ALL DELIVERIES shall be effected to the Institute for Stem Cell Science and Regenerative Medicine, National Centre for Biological Sciences, GKVK, Bellary Road, Bangalore-560065, whereas the foreign suppliers shall effect delivery up to Bangalore Airport, India by AIR only unless otherwise specified.

11. GUARANTEE/WARRANTY

The item/s covered under this tender shall be subject to a guarantee for trouble free performance, workmanship, material etc., fulfilling the specifications mentioned in this tender for min. **36 months** from the date of commissioning. If any defect is found in the material, workmanship or performance during the guarantee period the same may either be repaired/replaced by the supplier as the case may be free of charge. The guarantee period for replacement of parts or repair work shall be same as above. A guarantee certificate to this effect should be forwarded to us with your invoice. For order placed on foreign suppliers, the supplier shall specifically confirm that their Indian representatives, if any, will provide with after sales service and will attend to any repairs or technical problems that may arise.

12. PERFORMANCE B/G

The successful bidder shall have to execute a performance B/G in accordance with the guarantee/warranty for 3% value of the order. This B/G has to be executed on an appropriate value of stamp paper in terms of a bank guarantee drawn on any Nationalized Bank or first class Foreign Bank and shall remain valid till the completion of the Defect Liability period/warranty period, with **6 months** claim period.

13. TAXES AND DUTIES

The Centre shall deduct all taxes and duties, as applicable, from time to time from the bills payable.


14. PAYMENT TERMS

Payment Term: 90% against delivery; balance 10% on installation subject to receipt of PBG for 3% order value. The percentage of PBG will be determined based on the prevailing rules as of date of finalization of the tender

15. CLARIFICATIONS

After opening the bids, if it becomes necessary for the purchaser to seek clarifications from the bidders, the same will be sought from the bidders. In such an event, the bidders will furnish all technical information / clarifications to the purchaser to reach on or before the due date fixed for that purpose, indicating the Purchaser's tender reference. If the technical clarifications sought do not reach on or before the date fixed, the bids shall be summarily rejected without any further notice.





16. RISK CLAUSE

Notwithstanding the other terms therein, the Centre at its option will be entitled to terminate the contract and to avail from elsewhere; at the risk and cost of contractor; either the whole of the contract or any part which the contractor has failed to perform within the time stipulated or if the same performance is not available, the best and the nearest available substitute thereof. The contractor shall be liable for any loss which the Centre may sustain by reason of such risk contract in addition to penalty.

As per O.M No. F.No.6/18/2019-PPD, dt.23/07/2020, the following condition to be fulfilled and the bidder to submit the following declaration on their Letter head.

I. Any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is Registered with the Competent Authority.

II. "Bidder" (including the term 'tenderer', 'consultant' or 'service provider' in certain contexts) means any person or firm or company, including any member of a consortium or joint venture (that is an association of several persons, or firms or companies), every artificial juridical person not falling in any of the descriptions of bidders stated hereinbefore, including any agency branch or office controlled by such person, participating in a procurement process.

III. "Bidder from a country which shares a land border with India" for the purpose of this Order means:-

- a. An entity incorporated, established or registered in such a country; or
- b. A Subsidiary of an entity incorporated, established or registered in such a country; or
- c. An entity substantially controlled through entities incorporated, established or registered in such a country; or
- d. An entity whose beneficial owner is situated in such a country; or
- e. An Indian (or other) agent of such an entity; or
- f. A natural person who is a citizen of such a country; or
- g. A consortium or joint venture where any member of the consortium or joint venture falls under any of the above.

17. DISPUTE AND RESOLUTION

Any dispute or differences between the parties that cannot be settled by mutual discussion at appropriate levels shall be referred to the sole arbitration of the Director, INSTEM or his nominee and his decision in the matter shall be final and binding upon the parties to the dispute. The venue of arbitration proceedings shall be Bangalore. In respect of any matters pertaining to such arbitration, the courts of law in Bangalore will have exclusive jurisdiction. This Arbitration Agreement (hereinafter referred to as this "Agreement") relating to this Contract (hereinafter called the "Main Agreement" for this agreement) is made under the provisions of The Arbitration and Conciliation Act, 1996 as amended from time to time and the rules thereunder (hereinafter called The Arbitration Act).

18. OTHER TERMS

- a) If electrical/ AC/ technical works contract, appropriate license defining the required expertise from the approved Licensing authority.
- b) The Centre, will at its discretion ask for a solvency certificate from your bankers.

19. DEFINITION AND MEANINGS

In constructing these conditions, specifications, etc. in the tender document or the Annexures/Appendices the following words shall mean herein assigned definitions except where the subject context is otherwise stated.





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

PURCHASER:	Shall mean the Institute for Stem Cell Science and Regenerative Medicine
BID/TENDER:	Shall mean the proposal/document that the BIDDER submits in the requested and specified form or otherwise along with Annexures, Appendices, etc.
BIDDER/TENDERER:	Shall mean the firm/party who quotes against an enquiry.
CONTRACTOR/SUPPLIER:	Shall mean the party to whom a Work Order/Purchase Order is awarded to undertake all or a part of the work covered by this tender document as well as an amendment orders relating to this tender issued by the Purchaser and shall include his/their legal representative, assignee/s or successor/s.
CONTRACT:	Shall mean and include the articles of agreement, Declaration form, the general and special conditions, the Annexures, the Schedule of Quantities & rates and the specifications attached hereto and the drawings, if any.
ORDER VALUE:	Shall mean total value of the Purchase Order/Work order issued against this tender item including taxes, levies, etc.

20. TECHNICAL SPECIFICATIONS

See Annexure – 'A'

21. For imported items, final price after totaling the prices of the individual items has to be mentioned with estimated freight, insurance, agency commission charges, etc. The price shall be FOR CIP/CIF BANGALORE AIRPORT.
22. INSTEM reserves the right to split the quantities or reject one or more offers in full or part without any reasons. Therefore, INSTEM decision is final and binding.
23. The quantity mentioned in the tender may be decreased/increased while ordering. However, the price quoted should be firm irrespective of change in the quantity.
24. ACCEPTANCE OF TENDERS
INSTEM does not pledge itself to accept the lowest/ any tenders and reserves to itself the right to accept the whole or part of the tenders or a part of the quantity offered.
25. The firm may quote in Foreign Currency for direct import on CIP/CIF basis as well as in Rupees on FOR INSTEM basis for local supply. Please indicate insurance, freight and other charges separately for direct import.
26. Your Service Engineer should be fully trained to install the equipment and capable of maintaining the equipment during / after the warranty period.





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

27. LIQUIDATED DAMAGES:

The successful bidder will have to agree that in case the item is not supplied within the agreed delivery schedule and after a grace period of seven days, then Liquidated Damages (not in terms of penalty) will be imposed automatically and be deducted from their bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.

28. Micro and Small Enterprises (MSEs):

a. Micro and Small Enterprises (MSE) must, along with their offer, provide proof of their being registered as MSE (indicating the terminal validity date of their registration) for the item tendered, with any agency mentioned in the notification of the Ministry of Micro, Small and Medium Enterprises (Ministry of MSME)

b. The MSEs are exempted from payment of earnest money and tender fees subject to furnishing of relevant valid certificate for claiming exemption as per privilege rules of Government of India.

Since INSTEM is a Public Funded Research Organisation, registered under Government of India Notification No.51/96 – Customs and 10/97 – for Central Excise is applicable for the items listed in the notification.

29. Eligibility for Participation:

(i) Indigenous manufacturers and their authorized dealers are eligible and they will be considered as Local Supplier Class – I/II for price preference as per GOI Notification vide OM bearing No.F-45021/2/2017-PP(BE-II) dated 04th June, 2020 and again on 16th September 2020 issued by Department of Industry and Internal Trade (Public Procurement Section), Ministry of Commerce & Industry, Govt. of India.

(ii) In pursuant of the clarification vide GOI Notification vide OM bearing No.P-45021/102/2019-PP(BE-II) (E-29930) dated 26.11.2020 issued by Public Procurement Section, Department of Promotion of Industry and Internal Trade, Ministry of Commerce and Industry, GOI that bidders offering imported products will fall under the category of Non-Local Suppliers and they can't claim themselves as Class – I Local Supplier / Class – II Local Supplier by claiming profit, warehousing, marketing, logistic, freight etc., as Local Value addition.

(iii) Since it is Global Tender Enquiry (GTE) thus non supplier in terms of GOI Notification vide OM bearing No.F-45021/2/2017-PP(BE-II) dated 04th June, 2020 and again on 16th September 2020 issued by Department of Industry and Internal Trade (Public Procurement Section), Ministry of Commerce & Industry, Govt. of India are also eligible for participation in the tender.

(iv) In pursuant of the OM bearing No.F.N.6/18/2019-PPD dated 23.07.2020 issued by Public Procurement Div., Department of Expenditure, Ministry of Finance, Govt. of India the manufacturer/supplier are not eligible for participation if they are from such country which shares the land border with India. Offered items manufactured in such countries will not be considered.

The taxes (State, Central, Turnover tax, Works Contract Tax, etc.). Please specify which are applicable. The duties and other levies, freight, insurance shall be stated clearly and separately. Also please mention whether the same is included in the price/s quoted.





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

30. FORCE MAJEURE CLAUSE:

a) Force Majeure means an event beyond the control of the supplier and not involving the supplier's fault or negligence and which is not foreseeable. Such events may include, but are not restricted to the acts of the purchaser either in its sovereign or contractual capacity, wars or revolution, hostility, acts of the public enemy, civil commotion, sabotage, fires, floods, explosions, epidemics, quarantine restrictions, strikes, lock-outs, and freight embargoes. However, it should not be used by a party to effectively to escape liability for bad performance.


If there is delay in performance or other failures by the supplier to perform its obligation under its contract due to event of a Force Majeure, the supplier shall not be held responsible for such delays/failures.

If a force majeure situation arises, the supplier shall promptly notify the purchaser in writing of such conditions and the cause thereof within 21 days of occurrence of such event with reasonable evidence thereof. Unless otherwise directed by the purchaser in writing, the supplier shall continue to perform its obligations under the contract as far as reasonably practical, and shall seek all reasonable alternative means of performance not prevented by the force majeure event.

If the performance in whole or in part or any obligation under this contract is prevented or delayed by any reason of force majeure for a period exceeding 90 days, either party may at its option terminate the contract without any financial repercussion on either side.

b) There may be a Force Majeure situation affecting the institute only, under which the institute shall have equal rights to claim relief under such Force Majeure situations.

For and on behalf of
Institute for Stem Cell Science and Regenerative Medicine


Purchase Officer

GKVK Post, Bellary Road, Bangalore 560 065, India
Phone +91-80-23666341 /344/345/346 . Telefax +91-80-23636662
purchase@instem.res.in . www.instem.res.in





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

Declaration of Compliance of Order (Public Procurement No.1, 2 & 3) dtd 23-Jul-2020 & 24 Jul 2020 on Restrictions under Rule 144 (xi) of the General Financial Rules (GFRs), 2017

This declaration must form part of all tenders & it contains general information and serves as a declaration form for all bidders. (Before completing this declaration, bidders must study the General conditions. Definitions, Govt Directives applicable in respect of Public Procurement No.1, 2 & 3) dtd 23 Jul 2020 & 24 Jul 2020 on Restrictions under Rule 144 (xi) of the General Financial Rules (GFRs), 2017 & prescribed tender conditions).

DECLARATION BY AUTHORISED SIGNATORY OF THE FIRM

I, the undersigned, (full names), do hereby declare, in my capacity as of M/s. (name of bidder entity), that:

- 1) The facts contained herein are within my own personal knowledge.
- 2) I have read the Order (Public Procurement No.1, 2 & 3) dtd 23 Jul 2020 & 24 Jul 2020 on the subject of Restrictions under Rule 144 (xi) of the General Financial Rules (GFRs), 2017 regarding restrictions on procurement from a bidder of a country which shares a land border with India and comply to all the provisions of the Order.
- 3) I certify that M/s. (name of bidder entity) ***is not from such a country or, is from such a country (strike out whichever is not applicable)***, has been registered with the Competent Authority. I hereby certify that the SUPPLIER fulfills all requirements in this regard and is eligible to be considered. (Where applicable, evidence of valid registration by the Competent Authority is attached).
- 4) I understand that the submission of incorrect data and / or if certificate / declaration given by M/s. (name of bidder entity) is found to be false, this would be a ground for immediate termination and further legal action in accordance with law as per clause 12 of the Public Order on Restrictions under Rule 144 (xi) of the General Financial Rules (GFRs), 2017.

AUTHORISED SIGNATURE: _____ **DATE:** _____

Seal / Stamp of Bidder





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

Bid Security Declaration for Order Value > Rs.5 Lakh
(to be submitted on Company's Letter Head)

I/We the undersigned hereby declare that if we withdraw or modify the bids during the period of its validity, or if we are awarded the contract and fail to sign the contract, or to submit a performance security before the deadline defined in the request for bids document or fail to execute the contract, we will be suspended for the period of time specified in the request for bids document from being eligible to submit bids for contracts with the entity that invited the bids.

Name and Signature
of Authorized Signatory
and Company Seal



**ANNEXURE-A-SPECIFICATIONS****Light Sheet Microscope-Qty 01 No.****TECHNICAL SPECIFICATIONS:**

Si. No	Technical Requirement
1	Light sheet Microscope - Dual side illumination with multi-directional light sheet functionality, clearing prepared, two detection modules pco. edge, illumination and detection optics for moderate and high magnification for aqueous samples, and moderate magnification for large cleared samples, four laser lines 405nm, 488nm, 561nm, and 638nm, starter kit sample chamber & holder, sample chamber and holder sets for large cleared samples, automatic temperature control of the sample chamber, one highend PC for system control with real time controller.
2	Lightsheet detection optic 5x/0.16 - EC PlanNEOFLUAR 5x/0,16 , Detection Optic Adapter includes retaining ring for 5x Lightsheet. Lightsheet detection optics 5x/0,16 foc includes: - EC Plan-NEOFLUAR 5x/0,16, Focus ring to adapt to refractive index of Medium (clearing) Lightsheet detection optic 10x/0.5 includes: - W PlanApochromat 10x/0.5, - Detection Optic Adapter 20x Lightsheet detection optic 20x/1.0 includes: - W PlanAPO 20x/1,0 DIC - Detection Optic Adapter 20x Lightsheet Objective Clr Plan-Neofluar 20x/1.0 Corr nd=1.45 M32 85mm, Parfocal length 85mm, overall length 105mm. (FWD=5.6mm) VIS-NIR . Correction collar for compensation of spherical aberrations. Suitable for clearing methods with nd=1.45, in particular Focus Clear™
3	Illumination Optics Lightsheet 5x/0.1 foc Special optics to adapt to the Refractive Index of the medium. The 5x illumination optics is used in combination with detection objective lenses with a magnification of less than 20x. Illumination Optics Lightsheet 10x/0.2 foc Special optics to adapt to the Refractive Index of the medium. The 10x/0.2 illumination optics is used in combination with detection objective lenses with a magnification of 20x and higher
4	Chamber & Sample Holder Starter Kit Lightsheet: Sample holder disc, Sample holder stem universal, Sample holder stem for capillaries, Set of Capillary Sleeves for Lightsheet, Syringe Sample Holder Set Lightsheet, Sample Chamber Set water Lightsheet, Tubing to fill sample chamber (75cm and 100cm), 20 syringes to fill sample chamber (10 x 50 ml and 10 x 20 ml) , 10 syringes for sample mounting (1 ml) , Capillaries Consumables Kit Lightsheet. Sample chamber set Clear large Sample chamber set: Sample chamber set clearing large Lightsheet consisting of: - 1 Sample Chamber n=1.35-1.58, 20x - 1 Sample Chamber Dove Tail Slide Lightsheet - 2 drain connectors for the sample chamber of Lightsheet 7 to guide liquid excess to the drain - 8 Luer Lock connectors to fill or perfuse the sample chamber - 6 Blind plugs to close the sample chamber openings - 2 Blind plugs flat overflow seal - 8 Blind plugs flat inlet/outlet seal -5x Boxes of Coverslips (100 each box) (sample chamber windows) - 1 Sample Chamber Window tool sample chamber Clearing, - 8 Illumination adapter rings Clearing - 50 O-rings 17 mm for illumination adapter rings - insert for volume reduction for chamber Clear large Sample holder set for large specimen consisting of 1x





	<p>smart sample holder shaft with stem and disk , sample holder for large specimen , hexagon screwdriver Sample chamber set Clr 20x Sample chamber set Sample chamber set Clr 20x Sample chamber set clearing 20x Lightsheet 7 includes:</p> <ul style="list-style-type: none"> - 1 Sample Chamber n=1.35-1.58, 20x - 1 Sample Chamber Dove Tail Slide Lightsheet - 2 drain connectors for the sample chamber of Lightsheet to guide liquid excess to the drain - 8 Luer Lock connectors to fill or perfuse the sample chamber, 6 Blind plugs to close the sample chamber openings, 2 Blind plugs flat overflow seal - 8 Blind plugs flat inlet/outlet seal, 5x Boxes of Coverslips (100 each box) (sample chamber windows) - 1 Sample Chamber Window tool sample chamber Clearing, 8 Illuminationadapter rings Clearing - 50 O-rings 17 mm for illumination adapter rings - 20 O-rings 20 mm to seal the detection window of the sample chamber body - insert for volume reduction for chamber 20x - 1 tool for Correction collar, detection optics
5	<p>Temperature Module</p> <ul style="list-style-type: none"> - Basic stack module for the temperature control of up to 4 electrically heated S1 components - Power supply for further S1 modules - Eight allocatable parameter sets for control - Resolution: 0.1° C - Temperature accuracy: +/- 0.1°C - Temperature measuring channel for external calibration - CAN and USB connecting cables - 90V...260V AC/50...60Hz/max. 460VA - India - specific line cab <p>TempModule CZ-LSFM with accessories to drive the Peltier Block S. Includes:</p> <ul style="list-style-type: none"> - TempModule CZ-LSFM with electrical connection TempModule CZ-LSFM to Lightsheet System (Back side of Device) - electrical connector cable from TempModule S1 to TempModule CZ-LSFM - 2x tubing from TempModule CZ-LSFM to Lightsheet System (Back side of Device) - Peltier Block S with attached cables and connectors (2x cooling liquid, 1x electrical) - Stainless steel Temperature Sensor (CrNiMoTi alloy); by Mo higher resistance against salts in media - This Module requires a TempModul S1 to be operated
6	<p>CO2-Module Lightsheet (D) to incubate the Sample Chamber Lightsheet with CO2-enriched air. Includes:</p>





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

	<p>- CO2 Module, the necessary tubing to guide the gas into the sample chamber, sterile filter for clean air Requires a TempModule S1. Can be extended with Heating Device Humidity S1 to improve thermal equilibrium in the sample chamber.</p>
7	<p>Heating Device Humidity Solid State Laser 405nm, 20mW (D) Variability of wavelength from Laser to Laser: 405nm +/- 5nm Solid State Laser 488nm, 30mW (D) Variability of wavelength from Laser to Laser: 488nm +/- 5nm Solid State Laser 561nm, 20mW (D) Variability of wavelength from Laser to Laser: 561nm +/- 0.5nm Solid State Laser 638nm, 75mW (D) Variability of wavelength from Laser to Laser: 638nm +/- 5nm Laser Bench Rack Lightsheet including:</p> <ul style="list-style-type: none"> - 19" rack system with laser module and laser combiner module - 19" power supply module for the product Lightsheet - high precision connector to connect a laser fiber - laser fiber - AOTF for fast control of the laser illumination intensity (0-100%) and fast selection of wavelength - Laser module prepared with 6 beam combiners to connect up to 6 laser lines of the wavelengths (in nm): 405,445,488,515,561 and 638 - 1x USB interface to connect to PC - 1x LVDS interface to connect to Real Time Controller Imaging Systems - Power supply with wide input voltage range 100...125 V AC and 220...240 V AC ($\pm 10\%$) 50...60 Hz - India-specific power cable Dimensions: WxHxD (in cm) ca. 60x56x70, on wheels
8	Filter Module LBF 405/488/561/638 for Lightsheet
9	<p>Lightsheet Reflector Turret for Laser Blocking Filters Lightsheet Reflector Turret for Laser Blocking Filters with 5 free P & C Positions to accommodate the Filter Modules LBF of the Lightsheet and for their motorized, software-controlled positioning in the detection beam path (special P & C Filter Cubes for Lightsheet, including a notch filter to block the lines as well as an anti-reflection coated 2 mm compensation glass). The sixth position is occupied by a mandatory P & C with compensation glass which is required for light sheet adjustment. This Reflector Turret is required for system operation. A minimum of one turret must be ordered with each Lightsheet System</p>
10	<p>Camera pco.edge 4.2 sCMOS detection module, Liquid cooling system, Frame-grabber board. Lightsheet Reflector Turret for Emission Selection Can be used for "single" and "dual" detection module configurations: The turret has five positions which can be optionally equipped with Filter Modules. One position is occupied by a required "Filter Module Lightsheet empty with Mirror" The equipment and alignment of these positions must be done in the factory and are permanently fixed into the turret when the purchase order is placed. This Reflector Turret is required for system operation. A minimum of one turret must be ordered with each Lightsheet System</p>
11	<p>Filter Module Lightsheet --- DAPI-GFP Used for dual channel imaging in dual camera configurations. Mounted beam splitter SBS LP 490 and two emission filters BP 420-470 and BP 505-545 for optionally equipping one of the free positions in the Refl-Turret of the Lightsheet. The emission filters are suitable for simultaneous detection of the Fluorescence signal of DAPI and GFP or dyes with similar spectral characteristics.</p> <p>Filter Module Lightsheet --- GFP-Cy3 Used for dual channel imaging in dual camera configurations. Mounted beam splitter SBS LP 560 and two emission filters BP 505-545 and BP 575-615 for optionally equipping one of the free positions in the Refl-Turret of</p>





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

	<p>the Lightsheet. The emission filters are suitable for simultaneous detection of the Fluorescence signal of GFP and Cy3 or dyes with similar spectral characteristics.</p> <p>Filter Module Lightsheet --- GFP-mCherry Used for dual channel imaging in dual camera configurations. Mounted beam splitter SBS LP 560 and two emission filters BP 505-545 and LP 585 for optionally equipping one of the free positions in the Refl-Turret of the Lightsheet. The emission filters are suitable for simultaneous detection of the Fluorescence signal of GFP and mCherry or dyes with similar spectral characteristics.</p> <p>Filter Module Lightsheet --- GFP-DRAQ5 Used for dual channel imaging in dual camera configurations. Mounted beam splitter SBS LP 560 and two emission filters BP 505-545 and LP 660 for optionally equipping one of the free positions in the Refl-Turret of the Lightsheet. The emission filters are suitable for simultaneous detection of the fluorescence signal of GFP and DRAQ5 or dyes with similar spectral characteristics.</p> <p>Filter Module Lightsheet --- Cy3-DRAQ5 Used for dual channel imaging in dual camera configurations. Mounted beam splitter SBS LP 640 and two emission filters BP 575-615 and LP 660 for optionally equipping one of the free positions in the Refl-Turret of the Lightsheet. The emission filters are suitable for simultaneous detection of the Fluorescence signal of Cy3 and DRAQ5 or dyes with similar spectral characteristics</p>
12	<p>Language Pack English US (O) Microscopy Workstation Premium Intel Xeon multilingual (O) HP Z6 G4 Workstation</p> <ul style="list-style-type: none">- Chipset: Intel C622- Memory: max. 192 GB RAM- Memory Expansion Slots: 6 x DDR4, RDIMM, ECC, 2666 MHz Module- PCI Express Connectors: <p>2 x PCI Express Generation 3 x16 1 x PCI Express Generation 3 x8 3 x PCI Express Generation 3 x4</p> <ul style="list-style-type: none">- SSD: 1 x 512 GB M.2 NVMe- Hard Drives: 2 x 4 TB SATA 7200 rpm (configured as 4 TB RAID 1 hard drive)- Optical Storage: DVD +/- RW recorder for rewriteable media- Drive Bays: 4- I/O Controller: SATA 6.0 GB/s controller integrated with 6 ports- Network Controller: Dual integrated GbE LAN interface- I/O: 10 x USB 3.0 type A (4 x front, 6 x rear), PS/2 for keyboard and mouse 410203-9911-000 1 <p>System configuration 4 Windows 10 LTSB 2016 (hp Z6 R1) (O) Installation of Windows 10 Enterprise LTSB 2016 x64</p> <p>High end incl. Windows 10 license Installation of ZEN software for Lightsheet Network Adapter 2x10 GbE RJ45 (hp Z6) (O) Additional network adapter 2x10 GbE RJ45 (hp Z6) e.g. for connection of storage systems</p> <p>Real Time Controller Imaging Systems. Processor Intel Xeon Gold 6134 (hp Z6) (O) High performance processor configuration</p> <ul style="list-style-type: none">- Intel Xeon Gold 6134 (3.2 GHz, 24.75 MB cache, 8 cores).





	<p>Graphic Card NVIDIA Quadro P4000 8GB DP (O) Professional graphic card for Workstation Premium - 8 GB GDDR5 memory - PCI Express 3.0 Interface x16 - Hardware acceleration of DirectX 12, OpenGL 4.5 and CUDA - Support for up to four displays - Display outputs: 4 x DisplayPort 1.4 (DP) - DisplayPort support for 4 x 32" Display (3840 x 2160 Pixel). Memory 32 GB (1x32) DDR4-2666 MHz ECC registered RAM (O) Hard Drive Extension 12TB RAID10 (hp Z6) (O) Monitor TFT 32" HP Z32 (O)</p> <p>Visible diagonal: 80 cm, 16:9 Max. resolution: 3840 x 2160 pixel @ 60 Hz Brightness: 350 cd/m2 Contrast: 1300:1 Viewing angle: horizontal 178 degree, vertical 178 degree 3 x Video input: DisplayPort 1.2, mini DisplayPort 1.2, HDMI 2.0 3 x USB 3.0, 1 x USB 3.1 Type-C Power consumption: Maximum 185 W, Typical 70 W, Power Saving 0.5 W Dimensions with stand (H x W x D): 56.44 cm x 71.45 cm x 24.8 cm Weight: 12.2 kg (incl. stand) Ergonomic features: Tilt: -5°to +20°, swivel:90°, height adjustable 12 cm Certification and compliance: CE, CSA, GS, TÜV, FCC, TCO Operating voltage: 100-240 V, country-specific power cable. Storage and Data Analysis PC 55TB, RTX4000 (O) High Performance Data Analysis and Storage Workstation 2x (one redundant) high efficiency power supply - CPU: Intel P XEON E5-2630V4 2,2 GHz LGA2011 L3 25MB Box - Graphics: NVIDIA RTX4000 (PNY), 8 GB, Cuda capable, PCI Express 3.0x16 - Memory: 128GB included, max. 256 GB RAM; Memory slots: 16 x DIMM slots - Hard Drives: 6x HDD 12TB, RAID 5 configured to 55 TB data storage volume - 1 Hard Drive can be hot-swapped without loss of data (in case of a HDD failure) - 2x Solid State Drive 240 GB for pagefile and operating system - 10 Gbit Ethernet on motherboard and 10 GbE cable to connect with PC for system control (high speed data streaming) - LAN: 2 x 10 GbE - 5x USB 3.0, 4x USB 2.0 - Windows 10 IoT Enterprise 2019 LTSC</p>
13	System table, small, air damped, level regulated pressured air supply necessary, long and short side open, with breadboard, 900x750mm





14	<p>Processing Software Package</p> <p>Deconvolution Module for improvement of 3D image stacks with 3D deconvolution algorithms, methods for theoretical PSF. Significant performance gain when used with suitable Nvidia GPU. image processing under Win 7 / 10x64.</p> <p>Offline processing and analysis without microscope control, deliverable with an existing license.</p> <p>Includes the basic functionality, and the modules of the respective</p> <p>System license without image acquisition. The following modules are always included:</p> <ul style="list-style-type: none">-Module Measurement-Module Image Analysis-Module Extended Focus <p>Dongle / hardware license USB for ZEN (O)</p>
----	--



**INFORMATION TO TENDERERS****The Tender shall be evaluated under 2 (Two) Bid System****I Technical Bid****II Financial Bid****TECHNICAL SPECIFICATIONS & EVALUATION CRITERIA WITH MARKS FOR 2 PART
TENDER FOR "Light Sheet Microscope-Qty 01 No".**

TECHNICAL EVALUATION CRITERIA WITH MARKS		
Si. No	Technical Requirement	Marks
1	Light sheet Microscope - Dual side illumination with multi-directional light sheet functionality, clearing prepared, two detection modules pco.edge, illumination and detection optics for moderate and high magnification for aqueous samples, and moderate magnification for large cleared samples, four laser lines 405nm, 488nm, 561nm, and 638nm, starter kit sample chamber & holder, sample chamber and holder sets for large cleared samples, automatic temperature control of the sample chamber, one highend PC for system control with real time controller.	10
2	Lightsheet detection optic 5x/0.16 - EC PlanNEOFLUAR 5x/0,16 , Detection Optic Adapter includes retaining ring for 5x Lightsheet. Lightsheet detection optics 5x/0,16 foc includes: - EC Plan-NEOFLUAR 5x/0,16, Focus ring to adapt to refractive index of Medium (clearing) Lightsheet detection optic 10x/0.5 includes: - W PlanApochromat 10x/0.5, - Detection Optic Adapter 20x Lightsheet detection optic 20x/1.0 includes: - W PlanAPO 20x/1,0 DIC - Detection Optic Adapter 20x Lightsheet Objective Clr Plan-Neofluar 20x/1.0 Corr nd=1.45 M32 85mm, Parfocal length 85mm, overall length 105mm. (FWD=5.6mm) VIS-NIR . Correction collar for compensation of spherical aberrations. Suitable for clearing methods with nd=1.45, in particular FocusClearTM	10
3	Illumination Optics Lightsheet 5x/0.1 foc Special optics to adapt to the Refractive Index of the medium. The 5x illumination optics is used in combination with detection objective lenses with a magnification of less than 20x. Illumination Optics Lightsheet 10x/0.2 foc Special optics to adapt to the Refractive Index of the medium. The 10x/0.2 illumination optics is used in combination with detection objective lenses with a magnification of 20x and higher	10





4	<p>Chamber & Sample Holder Starter Kit Lightsheet: Sample holder disc, Sample holder stem universal, Sample holder stem for capillaries, Set of Capillary Sleeves for Lightsheet, Syringe Sample Holder Set Lightsheet, Sample Chamber Set water Lightsheet, Tubing to fill sample chamber (75cm and 100cm), 20 syringes to fill sample chamber (10 x 50 ml and 10 x 20 ml) , 10 syringes for sample mounting (1 ml) , Capillaries Consumables Kit Lightsheet.</p> <p>Sample chamber set Clear large Sample chamber set: Sample chamber set clearing large Lightsheet consisting of:</p> <ul style="list-style-type: none"> - 1 Sample Chamber n=1.35-1.58, 20x - 1 Sample Chamber Dove Tail Slide Lightsheet - 2 drain connectors for the sample chamber of Lightsheet 7 to guide liquid excess to the drain - 8 Luer Lock connectors to fill or perfuse the sample chamber - 6 Blind plugs to close the sample chamber openings - 2 Blind plugs flat overflow seal - 8 Blind plugs flat inlet/outlet seal -5x Boxes of Coverslips (100 each box) (sample chamber windows) - 1 Sample Chamber Window tool sample chamber Clearing, - 8 Illumination adapter rings Clearing - 50 O-rings 17 mm for illumination adapter rings - insert for volume reduction for chamber Clear large <p>Sample holder set for large specimen consisting of 1x smart sample holder shaft with stem and disk , sample holder for large specimen , hexagon screwdriver</p> <p>Sample chamber set Clr 20x Sample chamber set Sample chamber set Clr 20x Sample chamber set clearing 20x Lighsheet 7 includes:</p> <ul style="list-style-type: none"> - 1 Sample Chamber n=1.35-1.58, 20x - 1 Sample Chamber Dove Tail Slide Lightsheet - 2 drain connectors for the sample chamber of Lightsheet to guide liquid excess to the drain - 8 Luer Lock connectors to fill or perfuse the sample chamber, 6 Blind plugs to close the sample chamber openings, 2 Blind plugs flat overflow seal - 8 Blind plugs flat inlet/outlet seal, 5x Boxes of Coverslips (100 each box) (sample chamber windows) - 1 Sample Chamber Window tool sample chamber Clearing, 8 Illuminationadapter rings Clearing - 50 O-rings 17 mm for illumination adapter rings - 20 O-rings 20 mm to seal the detection window of the sample chamber body - insert for volume reduction for chamber 20x - 1 tool for Correction collar, detection optics 	10
5	<p>Temperature Module - Basic stack module for the temperature control of up to 4 electrically heated S1 components</p>	5





	<ul style="list-style-type: none"> - Power supply for further S1 modules - Eight allocatable parameter sets for control - Resolution: 0.1° C - Temperature accuracy: +/- 0.1°C - Temperature measuring channel for external calibration - CAN and USB connecting cables - 90V...260V AC/50...60Hz/max. 460VA - India - specific line cab <p>TempModule CZ-LSFM with accessories to drive the Peltier Block S. Includes:</p> <ul style="list-style-type: none"> - TempModule CZ-LSFM with electrical connection TempModule CZ-LSFM to Lightsheet System (Back side of Device) - electrical connector cable from TempModule S1 to TempModule CZ-LSFM - 2x tubing from TempModule CZ-LSFM to Lightsheet System (Back side of Device) - Peltier Block S with attached cables and connectors (2x cooling liquid, 1x electrical) - Stainless steel Temperature Sensor (CrNiMoTi alloy); by Mo higher resistance against salts in media - This Module requires a TempModul S1 to be operated 	
6	<p>CO2-Module Lightsheet (D) to incubate the Sample Chamber Lightsheet with CO2-enriched air.</p> <p>Includes:</p> <ul style="list-style-type: none"> - CO2 Module , the necessary tubing to guide the gas into the sample chamber , sterile filter for clean air <p>Requires a TempModule S1. Can be extended with Heating Device Humidity S1 to improve thermal equilibrium in the sample chamber.</p>	5
7	<p>Heating Device Humidity</p> <p>Solid State Laser 405nm, 20mW (D) Variability of wavelength from Laser to Laser: 405nm +/- 5nm</p> <p>Solid State Laser 488nm, 30mW (D) Variability of wavelength from Laser to Laser: 488nm +/- 5nm</p> <p>Solid State Laser 561nm, 20mW (D) Variability of wavelength from Laser to Laser: 561nm +/- 0.5nm</p> <p>Solid State Laser 638nm, 75mW (D) Variability of wavelength from Laser to Laser: 638nm +/- 5nm</p> <p>Laser Bench Rack Lightsheet including:</p> <ul style="list-style-type: none"> - 19" rack system with laser module and laser combiner module - 19" power supply module for the product Lightsheet - high precision connector to connect a laser fiber - laser fiber - AOTF for fast control of the laser illumination intensity (0-100%) and fast selection of wavelength - Laser module prepared with 6 beam combiners to 	5





	connect up to 6 laser lines of the wavelengths (in nm): 405,445,488,515,561 and 638 - 1x USB interface to connect to PC - 1x LVDS interface to connect to Real Time Controller Imaging Systems - Power supply with wide input voltage range 100...125 V AC and 220...240 V AC ($\pm 10\%$) 50...60 Hz - India-specific power cable Dimensions: WxHxD (in cm) ca. 60x56x70, on wheels	
8	Filter Module LBF 405/488/561/638 for Lightsheet	5
9	Lightsheet Reflector Turret for Laser Blocking Filters Lightsheet Reflector Turret for Laser Blocking Filters with 5 free P & C Positions to accommodate the Filter Modules LBF of the Lightsheet and for their motorized, software-controlled positioning in the detection beam path (special P & C Filter Cubes for Lightsheet, including a notch filter to block the lines as well as an anti-reflection coated 2 mm compensation glass). The sixth position is occupied by a mandatory P & C with compensation glass which is required for light sheet adjustment. This Reflector Turret is required for system operation. A minimum of one turret must be ordered with each Lightsheet System	5
10	Camera pco.edge 4.2 sCMOS detection module, Liquid cooling system, Frame-grabber board. Lightsheet Reflector Turret for Emission Selection Can be used for "single" and "dual" detection module configurations: The turret has five positions which can be optionally equipped with Filter Modules. One position is occupied by a required "Filter Module Lightsheet empty with Mirror" The equipment and alignment of these positions must be done in the factory and are permanently fixed into the turret when the purchase order is placed. This Reflector Turret is required for system operation. A minimum of one turret must be ordered with each Lightsheet System	10
11	Filter Module Lightsheet --- DAPI-GFP Used for dual channel imaging in dual camera configurations. Mounted beam splitter SBS LP 490 and two emission filters BP 420-470 and BP 505-545 for optionally equipping one of the free positions in the Refl-Turret of the Lightsheet. The emission filters are suitable for simultaneous detection of the Fluorescence signal of DAPI and GFP or dyes with similar spectral characteristics. Filter Module Lightsheet --- GFP-Cy3 Used for dual channel imaging in dual camera configurations. Mounted beam splitter SBS LP 560 and two emission filters BP 505-545 and BP 575-615 for optionally equipping one of the free positions in the Refl-Turret of the Lightsheet. The emission filters are suitable for simultaneous	10





	<p>detection of the Fluorescence signal of GFP and Cy3 or dyes with similar spectral characteristics.</p> <p>Filter Module Lightsheet --- GFP-mCherry Used for dual channel imaging in dual camera configurations. Mounted beam splitter SBS LP 560 and two emission filters BP 505-545 and LP 585 for optionally equipping one of the free positions in the Refl-Turret of the Lightsheet. The emission filters are suitable for simultaneous detection of the Fluorescence signal of GFP and mCherry or dyes with similar spectral characteristics.</p> <p>Filter Module Lightsheet --- GFP-DRAQ5 Used for dual channel imaging in dual camera configurations. Mounted beam splitter SBS LP 560 and two emission filters BP 505-545 and LP 660 for optionally equipping one of the free positions in the Refl-Turret of the Lightsheet. The emission filters are suitable for simultaneous detection of the fluorescence signal of GFP and DRAQ5 or dyes with similar spectral characteristics.</p> <p>Filter Module Lightsheet --- Cy3-DRAQ5 Used for dual channel imaging in dual camera configurations. Mounted beam splitter SBS LP 640 and two emission filters BP 575-615 and LP 660 for optionally equipping one of the free positions in the Refl-Turret of the Lightsheet. The emission filters are suitable for simultaneous detection of the Fluorescence signal of Cy3 and DRAQ5 or dyes with similar spectral characteristics</p>	
12	<p>Language Pack English US (O) Microscopy Workstation Premium Intel Xeon multilingual (O) HP Z6 G4 Workstation - Chipset: Intel C622 - Memory: max. 192 GB RAM - Memory Expansion Slots: 6 x DDR4, RDIMM, ECC, 2666 MHz Module - PCI Express Connectors: 2 x PCI Express Generation 3 x16 1 x PCI Express Generation 3 x8 3 x PCI Express Generation 3 x4 - SSD: 1 x 512 GB M.2 NVMe - Hard Drives: 2 x 4 TB SATA 7200 rpm (configured as 4 TB RAID 1 hard drive) - Optical Storage: DVD +/- RW recorder for rewriteable media - Drive Bays: 4 - I/O Controller: SATA 6.0 GB/s controller integrated with 6 ports - Network Controller: Dual integrated GbE LAN interface</p>	5





<p>- I/O: 10 x USB 3.0 type A (4 x front, 6 x rear), PS/2 for keyboard and mouse 410203-9911-000 1 System configuration 4 Windows 10 LTSB 2016 (hp Z6 R1) (O) Installation of Windows 10 Enterprise LTSB 2016 x64</p> <p>High end incl. Windows 10 license Installation of ZEN software for Lightsheet Network Adapter 2x10 GbE RJ45 (hp Z6) (O) Additional network adapter 2x10 GbE RJ45 (hp Z6) e.g. for connection of storage systems</p> <p>Real Time Controller Imaging Systems. Processor Intel Xeon Gold 6134 (hp Z6) (O) High performance processor configuration - Intel Xeon Gold 6134 (3.2 GHz, 24.75 MB cache, 8 cores). Graphic Card NVIDIA Quadro P4000 8GB DP (O) Professional graphic card for Workstation Premium - 8 GB GDDR5 memory - PCI Express 3.0 Interface x16 - Hardware acceleration of DirectX 12, OpenGL 4.5 and CUDA - Support for up to four displays - Display outputs: 4 x DisplayPort 1.4 (DP) - DisplayPort support for 4 x 32" Display (3840 x 2160 Pixel). Memory 32 GB (1x32) DDR4-2666 MHz ECC registered RAM (O) Hard Drive Extension 12TB RAID10 (hp Z6) (O) Monitor TFT 32" HP Z32 (O)</p> <p>Visible diagonal: 80 cm, 16:9 Max. resolution: 3840 x 2160 pixel @ 60 Hz Brightness: 350 cd/m2 Contrast: 1300:1 Viewing angle: horizontal 178 degree, vertical 178 degree 3 x Video input: DisplayPort 1.2, mini DisplayPort 1.2, HDMI 2.0 3 x USB 3.0, 1 x USB 3.1 Type-C Power consumption: Maximum 185 W, Typical 70 W, Power Saving 0.5 W Dimensions with stand (H x W x D): 56.44 cm x 71.45 cm x 24.8 cm Weight: 12.2 kg (incl. stand) Ergonomic features: Tilt: -5°to +20°, swivel:90°, height adjustable 12 cm Certification and compliance: CE, CSA, GS, TÜV, FCC, TCO Operating voltage: 100-240 V, country-specific power cable. Storage and Data Analysis PC 55TB, RTX4000 (O) High Performance Data Analysis and Storage Workstation</p>	
--	--





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

	<p>2x (one redundant) high efficiency power supply</p> <ul style="list-style-type: none"> - CPU: Intel P XEON E5-2630V4 2,2 GHz LGA2011 L3 25MB Box - Graphics: NVIDIA RTX4000 (PNY), 8 GB, Cuda capable, PCI Express 3.0x16 - Memory: 128GB included, max. 256 GB RAM; Memory slots: 16 x DIMM slots - Hard Drives: 6x HDD 12TB, RAID 5 configured to 55 TB data storage volume - 1 Hard Drive can be hot-swapped without loss of data (in case of a HDD failure) - 2x Solid State Drive 240 GB for pagefile and operating system - 10 Gbit Ethernet on motherboard and 10 GbE cable to connect with PC for system control (high speed data streaming) - LAN: 2 x 10 GbE - 5x USB 3.0, 4x USB 2.0 - Windows 10 IoT Enterprise 2019 LTSC 	
13	<p>System table, small, air damped, level regulated pressured air supply necessary, long and short side open, with breadboard, 900x750mm</p>	5
14	<p>Processing Software Package Deconvolution Module for improvement of 3D image stacks with 3D deconvolution algorithms, methods for theoretical PSF. Significant performance gain when used with suitable Nvidia GPU. image processing under Win 7 / 10x64. Offline processing and analysis without microscope control, deliverable with an existing license. Includes the basic functionality, and the modules of the respective System license without image acquisition. The following modules are always included: -Module Measurement -Module Image Analysis -Module Extended Focus Dongle / hardware license USB for ZEN (O)</p>	5
	Total	100
<p>Evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.</p>		





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

ANNEXURE – B
SCHEDULE OF EXPERIENCE

Please furnish list of major supplies/works undertaken in relevant to this tender during the past two years.

Sl. No.	Name of the company with full address	Name of project	Purchase Order No. & date	Brief description	Value (Rs.)
Signature					
Name					
Designation					
Name of company					
Date					
Seal of the company					





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

ANNEXURE - C

SCHEDULE OF DEVIATION FROM SPECIFICATIONS/CONDITIONS

All deviations from the specifications/conditions shall be filled in by the bidder in this schedule.

--

The bidder hereby certifies that the above mentioned are the only deviations from Technical Specification of this tender. [State NIL if no deviation is envisaged].

Signature	
Name	
Designation	
Name of the company	
Date	
Seal of the company	





inStem

Ref: INS/F-5537/2021-2022(Y)

Tender Notice No.003/2021-2022

COVER – II PRICE BID

Light Sheet Microscope-Qty 01 No

Sl. No.	Item Description	Qty	Unit Price	Total Price

Please enter Unit Price, Total Price and Grand Total in figures as well as words. Unit price and Total Price may be entered below each item.



