



Ref: INS/F-5764/2017-2018(Y)

Date : 23/11/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Environmental Chamber (Specifications Attached)	any	1.00	Nos.

Remarks: The Tenders to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".
2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
3. Country of origion
4. Place of Shipment
5. Name, address,telephone number and fax numbers of your bankers.
6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
- 7.**Please Indicate frieght and insurance charges seperately.**
8. Please Indicate the Warranty Period and Delivery Period.
9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **07/12/2017 till 5.30 p.m.**

Important Note

10. Shipping by AIRFREIGHT ONLY-CARGO MODE
11. Port of Entry for Customs clearance is -BANGALORE
12. Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5764/2017-2018(Y)

Date : 23/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advolerum duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in blue ink, appearing to read 'Yesu R'.

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



Institute for Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Dept of Biotechnology, Govt. of India.



INS/F-5764/2017-2018(Y)

Specification Of Environmental Chamber

1. System should be vertical standing and should have working volume of 800 Litres.
2. The system should be a research grade and suitable for continuous operation.
3. The system should have the temperature range of 10°C to 50°C.
4. The system should have temperature accuracy of $\pm 0.5^\circ\text{C}$ with Lights off and $\pm 2^\circ\text{C}$ with lights on.
5. The system should have relative humidity range from 40% to 100% with accuracy of $\pm 1\%$.
6. The system should have ultrasonic humidifier with electronic RH sensor.
7. The system should have controls for humidification and dehumidification.
8. The system should have microprocessor based temperature and humidity controllers.
9. The system should have alarms for temperature, humidity, door open.
10. The system should have option for data logging, alarm output for remote monitoring.
11. The system should have option for temperature, humidity calibration.
12. The system should have cool white fluorescent light with real time programmable intensity control.
13. The system should have programmes for controlling humidity, temperature, light intensity and time.
14. The system should have diffused air flow inside the chamber.
15. The system should have temperature safety limit control.
16. The system should have LCD/LED display with keypad for control functions.
17. The system should have 4-5 shelves made up of stainless steel with insulation.
18. The interior of the system should be made up of stainless steel.
19. The system exterior should be powder coated.
20. The system should be CFC, HCFC free.
21. The system should have caster wheels with lock.
22. The system should work on 230V, 50Hz
23. The system should be supplied with all the accessories required to function.
24. The system should be CE/ISO certified.
25. The system should have warranty of 3 years.

