Institute For Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Department of Biotechnology, Government Of India



Ref:

INS/F-5767/2017-2018(Y)

Date: 20/11/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing insurance, estimated airfreight charges handling charges etc CIE/CIP unto inStem Stores for the following materials

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	NOM
1		Bacterial incubator (As per Specification Attached)	any	7.00	Nos.

Remarks: The Tender to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Please also indicate specifically the following:

- 1. The bids / quotation addressed to "THE HEAD-PURCHASE" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected, if not addressed to "THE HEAD-PURCHASE".
- 2. EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)
- 3. Country of origion
- 4. Place of Shipment
- 5. Name, address, telephone number and fax numbers of your bankers.
- 6. Your Account number and swift number with the bank (required for payment purposes if the order materialises on your firm).
- 7. Please Indicate frieght and insurance charges seperately.
- 8. Please Indicate the Warranty Period and Delivery Period.
- 9.Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by 04/12/2017 till 5:30pm

Important Note

- 10.Shipping by AIRFREIGHT ONLY-CARGO MODE
- 11.Port of Entry for Customs clearance is -BANGALORE
- 12.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo, Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5767/2017-2018(Y)

Date: 20/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advolerum duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell Bidlogy and Regenerative Medicine

GKVK, Bellary Road, Bangalore-560065, DIA

Phone No.:

91-80-23666343/344/345/346

Email Id:

purchase@ncbs.res.in

Fax: 91

91-80-23636662

Website:

www,ncbs.res,in





REF: INS/F-5767/2017-2018(Y)

Specification of Bacterial Incubator

- 1. General: microprocessor controlled Incubator It should have LED Display with easy icons and user-friendly control system.
- 2. The system should have the capacity of 50 liters \pm 5%
- 3. The system should have temperature range from ambient+5 to70°C.
- 4. The system should have anoption to mute the audible alarms
- 5. The fluctuation/variation at 37°C should not be more the 0.1°C /0.5°C
- 6. The system should have facility to calibrate the temperature via keypad
- 7. The system should be supplied with 4nos of adjustable shelves
- 8. The system should have the Air circulation by a fan to maintain temperature uniformity.
- 9. The system should have the Corrosion resistance stainless steel interior and epoxy-powder coated exterior.
- 10.Electronic and mechanical components should be enclosed, protected from accidental spills.
- 11.It should have automatic restart after power interruption with nonvolatile memory
- 12. The system should operate at 230V, 50 Hz AC power.
- 13. The system should be supplied with all the accessories required to function.
- 14. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required).
- 15. The system should be supplied with 3 years warranty.
- 16. The System should be CE/ISO Approved

