

C

Institute For Stem Cell Biology and Regenerative Medicine
Autonomous Institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5768/2017-2018(Y)

Date : 20/11/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Incubator Shaker (As per specification attached)	Any	3.00	Nos.

Remarks: The Tender to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".
2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
3. Country of origion
4. Place of Shipment
5. Name, address,telephone number and fax numbers of your bankers.
6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
7. **Please Indicate frieght and insurance charges seperately.**
8. Please Indicate the Warranty Period and Delivery Period.
9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **04/12/2017 till 5:30pm**

Important Note

10. Shipping by AIRFREIGHT ONLY-CARGO MODE
11. Port of Entry for Customs clearance is -BANGALORE
12. Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine

Autonomous Institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5768/2017-2018(Y)

Date : 20/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advalorem duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to be "Yes R".

Yes R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in





REF: INS/F-5768/2017-2018(Y)

Specification of Incubator Shaker

1. System should be microprocessor controlled Incubator Shaker It should have with touch screen with easy icons and user-friendly control system.
2. The system should have the capacity of 80 liters \pm 2%
3. The system should be equipped with large, easy to read display screen.
4. The system should have temperature range from $+4^{\circ}\text{C}$ to 80°C .
5. The system speed should be 20-400rpm \pm 1rpm
6. The fluctuation/variation at 37°C should not be more the 0.1°C / 0.5°C
7. The system should have the option to mute the audible alarms
8. Air circulation by a fan to maintain temperature uniformity.
9. It should have timer of up to 1 min to 999 Hrs.
10. The system should have temperature calibration feature.
11. The system should have the Corrosion resistance stainless steel interior and epoxy-powder coated exterior.
12. The system should have Temperature and shaking speed deviation alarm.
13. The system should be supplied with 3 no's of adjustable shelves
14. The system should have Built-in RS-232 port and USB port for external control and data collection.
15. The system should have Smooth start and smooth stop function prevents chemical spills from flasks or test tubes and Easy-access drain system to clean up the spills.
16. The system should have programmable controller to automate parameter changes on a timed basis.



Institute For Stem Cell Biology And Regenerative Medicine

17. The system should be quoted with Universal Platform having capacity to holds various size of flask sizes 50ml, 100ml, 200ml, 250 ml, 500 ml, 1000 ml clamps each 2 nos and test tube racks with the main offer.
18. Electronic and mechanical components should be enclosed, protected from accidental spills.
19. The system should supply with universal platform with 14"wide X 14" depth
20. The system should have automatic restart function in case of power failure and after door opening and closing
21. The system should have acceleration/deceleration circuit to prevent sudden start and stop for delicate sample.
22. The incubator shaker should be quoted with stackable kit with the main offer.
23. The system should operate at 230V, 50 Hz AC power
24. The system should be supplied with Retractable foot caster for easy mobility during installation or relocation.
25. It should have open lid cut out switch stop agitation when chamber is open.
26. The system should be supplied with all the accessories required to function.
27. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required).
28. The system should be supplied with 3 years warranty.
29. The System should be CE/ISO Approved

