

Ref: INS/F-5771/2017-2018(Y)

Date : 30/11/2017

### ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Water bath with shaking and circulation (Specifications Attached)	any	10.00	Nos.
2		Vortex Mixer (Specifications Attached)	any	10.00	Nos.

**Remarks: The Tenders to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.**

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".

2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**

3. Country of origion

4. Place of Shipment

5. Name, address,telephone number and fax numbers of your bankers.

6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).

7. **Please Indicate freight and insurance charges seperately.**

8. Please Indicate the Warranty Period and Delivery Period.

9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **11/12/2017 till 5.30 p.m.**

**Important Note**

10. Shipping by AIRFREIGHT ONLY-CARGO MODE

11. Port of Entry for Customs clearance is -BANGALORE

12. Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



**Institute For Stem Cell Biology and Regenerative Medicine**  
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5771/2017-2018(Y)

Date : 30/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advolerum duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell  
Biology and Regenerative Medicine

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



Ref No: INS/F-5771/2017-2018(Y)

## **Shaking/ circulating Waterbath**

The system should have the capacity of 20 liters

The system should be microprocessor controlled and It should have the display with easy icons and user-friendly control system.

The water bath should have shaking option and should be controllable with a separate controller with speed control.

The system should have the Temperature range Ambient +5°C to 99.9°C

Temperature stability should be +/- 0.2C

The system should be supplied with the lid, Metal blocks, metal grill

The system should have the overheat protector

The SS Tank should be seamless stainless steel and outside body should be powder coated

The system should have the LED/LCD Display

Electrical supply 230V, 50Hz

The system should be CE/ISO Certified

The system should be supplied with all the accessories required to function

### **Specifications of vortex mixer**

The system should have the Speed Range 200-2000 rpm

The system should have Operating Modes Touch and Continuous.

The system should operate in temperature Range of ambient +4°C to 65°C

The system should have 3-way power switch,

The system should have the LED display for time/speed.

The system should operate on 230V, 50Hz,

The system should be CE/ISO Certified

The system should be supplied with all the accessories required to function.

The system should have the warranty for 3 years

