

Ref: INS/F-5783/2017-2018(Y)

Date : 20/11/2017

**ENQUIRY**

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Circulating water bath (Specification Attached)	any	1.00	Nos.

**Remarks:** The Tender to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".
2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
3. Country of origion
4. Place of Shipment
5. Name, address,telephone number and fax numbers of your bankers.
6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
- 7.**Please Indicate freight and insurance charges seperately.**
8. Please Indicate the Warranty Period and Delivery Period.
9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **04/12/2017 till 5:30pm**

Important Note

10. Shipping by AIRFREIGHT ONLY-CARGO MODE
11. Port of Entry for Customs clearance is -BANGALORE
12. Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



**Institute For Stem Cell Biology and Regenerative Medicine**

Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5783/2017-2018(Y)

Date : 20/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advalorem duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell  
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to be 'Kesu R'.

Kesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in





REF: INS/F-5783/2017-2018(Y)

### **Specification for Circulating Water bath**

1. The system should have the capacity of 20 liters
2. The system should be microprocessor controlled and It should have LED/LCD display with easy icons and user-friendly control system.
3. The system should have the temperature range from -10 °C to 50 °C
4. The system should have the Temperature stability +/- 0.2C
5. The system should have the maximum flow rate of 14L/min
6. The system should have the maximum pressure 1 bar/14psi
7. The system should be supplied with the lid, Metal blocks, metal grill
8. The system should have the overheat protector, low/high voltage protection, surge protection to minimize potential damage from voltage spikes.
9. The water bath should have option to switch off the circulating motor.
10. The system should be equipped with buzzer and visual alarm for power failure, and temperature deviation against the setup temperature Preferably there should be an option to mute the audible alarms
11. The mains switch should be protected in case of spillage of water.
12. It should have automatic restart after power interruption with non-volatile memory.
13. The SS Tank should be seamless stainless steel and outside body should be powder coated
14. The system should have the noise level of <60 dBA.
15. The system should be supplied with all the accessories required to function.
16. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required).
17. The system should be supplied with 3 years warranty as a part of the offer.
18. The System should be CE/ISO Approved

