



Ref: INS/F-5792/2017-2018(Y)

Date : 09/12/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Micro-pipette Beveller (As per specification attached)	Any	1.00	No.

Remarks: The Tenders can also to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".
2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
3. Country of origion
4. Place of Shipment
5. Name, address,telephone number and fax numbers of your bankers.
6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
- 7.**Please Indicate frieght and insurance charges seperately.**
- 8.Please Indicate the Warranty Period and Delivery Period.
- 9.Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **14/12/2017 till 5.30pm**

Important Note

- 10.Shipping by AIRFREIGHT ONLY-CARGO MODE
- 11.Port of Entry for Customs clearance is -BANGALORE
- 12.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5792/2017-2018(Y)

Date : 09/12/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advalorem duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

A handwritten signature in black ink, appearing to be 'Yesu R'.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

Yesu-R

GKVK, Bellary Road, Bangalore-560065,INDIA


Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in





INS/F-5792/2017-2018(Y)

Specification for Micropipette Beveller

- 1) System should be precision bevelling of micropipette tips between 0.1 and 100 μ m.
- 2) System should have unique abrasive plate drive and vibration free for greater control of the bevelling.
- 3) System should consists of a stationary pedestal, optically flat to a half wave , surface mounted on a heavy baseplate. The material of construction should be stainless steel.
- 4) System should have Bevelling Range 0.1 μ m through 100 μ m finished electrodes depending on abrasive plate used.
- 5) System should have Synchronous clock motor insures stable rotation rate.
- 6) System should have Grinding Speed greater than 50 RPM \pm 20%
- 7) Grinding Surface Variation less than 1.0 μ m
- 8) System should have adjustable Bevelling Angle Range 5-90 degrees
- 9) System should have Robust micromanipulator controls bevel angle and advancement.
- 10) System should have 2-axis micromanipulator with course drive: 0.075in / dial revolution and fine drive: 0.0004in / dial revolution.
- 11) System should have 80X stereo microscope for monitoring the bevelling process.
- 12) System should have LED lamp which enhances the bevelling operation by providing sharp illumination of the abrasive plate and pipette.
- 13) System should have supply of 230V, 50HZ
- 14) System should be CE/ISO certified
- 15) System should have the warranty for 3years.
- 16) System should be supplied with all accessories required for function.

