

Ref: INS/L-5146/2019-2020(Y)

Date : 17/05/2019

ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following materialises.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Vacuum Concentrators Volume (Specifications Attached)		1.00	No.

Remarks : The Price quoted against this RFQ should be extended to Bangalore Life Science Cluster (BLiSc) of NCBS, InStem and C-CAMP for placing repeat order as per norms, by any one or all institutes of Bangalore Life Science Cluster (BLiSc) **inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification. **

1. The bids shall be enclosed in an envelope, and due date sealed duly marked "Tender for _____" Ref No : _____. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission Date and Time and inStem will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS
03/06/2019 till 5.30 p.m.**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & Warranty period.

7. Delivery to be made to our Stores. Please indicate charges, if any extra. Transit Insurance should be done upto inStem Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment : within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs Duty (Except ad valorem duty of 5% + 2% cess and CVD @ 4% vide Notification No. 51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.

If the items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.



Ref: INS/L-5146/2019-2020(Y)

Date : 17/05/2019

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The Decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Institute For Stem Cell
Biology and Regenerative Medicine



Yesu R



GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



INS/L-5146/2019-2020(Y)

Specification for Vacuum Concentrator

1. Vacuum concentrator that supports the rapid and safe concentration and drying-down of samples, including nucleic acids, PCR preps and synthetic oligos.
2. External post trap to neutralize am

Materials of Construction	
Chamber	Teflon coated aluminum
Chamber Seal	Silicone rubber
Pump Heads	2-headed pump with ETFE diaphragms
Cover	Transparent acrylic lid with safety lock
Operational Requirements	
Voltage Input/Frequency/Power	
	Should work in the Indian electrical system
Typical Rotor Spin Speed	1300 RPM (Max)
Motor	Maintenance-free induction motor
Performance Properties	
Noise	<65 dBA
Vacuum Pump	ETFE oil-free diaphragm pump
Capacity	30 L/min
Vacuum Control (Ultimate Vacuum Level)	<10 Torr (13 mbar, 1.3 kPa)
Temperature Range	Ambient* to 65 °C
Display/User Interface	
Programs	3 selectable programs by application should be modified and stored, Auto and manual run selector
Control Panel	LED read out of temperature, run time and heat time
Temperature Set Point	Ambient, 35 - 65 °C in 5 °C increments
Timer Range (Heat and Run)	1 minute to 9 hours and 59 minutes, continuous
Modes of Operation	Auto, Manual and Pre-heat temperature
Compliance	
Compliance	UL listed to U.S./Canadian requirements. CE marked.

