

Ref: INS/L-5387/2020-2021(Y)

Date : 20/10/2020

ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Laminar air flow bench with stand, Vertical type (Specifications attached)		1.00	Nos.

Note: If the items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for _____" Ref No : _____. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

2. **DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS
04/11/2020 till 5.30 p.m.**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & Warranty period.

7. Delivery to be made to our Stores. Please indicate charges, if any extra. Transit Insurance should be done upto inStem Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, Please state so and then offer alternative to quality/Specifications.

9. Payment : within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs Duty (Except ad valorem duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.



Ref: INS/L-5387/2020-2021(Y)


Date : 20/10/2020

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Institute For Stem Cell
Science and Regenerative Medicine



Yesu R



GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in



Ref No: INS/L-5387/2020-2021(Y)

Technical Specification for Laminar Air flow

- 1) Laminar Cabinets Inner Chamber & Outer Chamber made of stainless steel high efficiency particular air filters, to achieve the air purification upto 0.3 Microns.
- 2) Size: Overall: 950L ×680D ×2100H mm (approx) work area: 870L ×630D × 610H mm (approx)
- 3) Blower fitted with ¼ HP Motor, with RPM 1200 to 1400.
- 4) Pre-filters made of high grade nylon Net fixed in S.S. frame for first Stage air purification, through blower system.
- 5) Closed Inner Chamber fitted with HEPA having very accurate performance rate of air filtration, rated 99.99%, resulting in ceasing all air bore molecule of particle up to 0.3 micron in working Area of Laminar Bench.
- 6) Working area of Laminar Airflow Cabinets illuminated by fluorescent light; cabinets operated at 230V. Single Phase 50Hz. AC Supply.
- 7) Fitted with UV Germicidal lamp for sterilization.
- 8) Fitted with Acrylic Front Door sliding type
- 9) Fitted with Manometer for Measurement of HEPA Filters Choking system.
- 10) Fitted with Cock for Gas Connection.
- 11) Should have technical support in Bangalore.



