

Ref: INS/L-5501/2017-2018(Y)

Date : 20/09/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Liquid Nitrogen dewars 4 liters (As per the specification attached)	Any	2.00	Nos.
2		Liquid Nitrogen dewars 10 liters (As per the specification attached)	Any	2.00	Nos.
3		Liquid Nitrogen dewars 25 liters (As per the specification attached)	Any	2.00	Nos.

Remarks :

Note :Quotation in USD, GBP, YEN, EURO, CND, SGD, ect will be acceptable.

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for _____" Ref No : _____. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS
09/10/2017 till 5.30 p.m.**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5.All duties,taxes,surcharge and cess as currently applicable must be sated in your quotation,seperately.Otherwise your quote is liable to be rejected.

6.Your quotation should indicate delivery period & Warranty period.

7.Delivery to be made to our Stores.Please indicate charges,if any extra.Transit Insurance should be done upto inStem Stores.

8.If you are unable to supply the quality,specifications or brand as mentioned in our enquiry,Please state so and then offer alternative to quality/Specifications.

9.Payment :within one month after delivery & acceptance/satisfactory installation.

10.Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11.Since we are a public funded research institution,we are exempted from paying Customs Duty(Except advolerum duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments,technical instruments,equipments(including computers),their accessories,spares,consumables and software.Hence,please offer your prices taking this option into consideration.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5501/2017-2018(Y)

Date : 20/09/2017

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in





InStem

Institute For Stem Cell Biology And Regenerative Medicine

Ref No: INS/L-5501/2017-2018(Y)

Specification for Liquid Nitrogen canister Dewar 4 liters

- 1) System should have capacity of 4 liters with wide neck opening for storing the samples.
- 2) System should be compatible for the storage of Biological materials with excellent holding time.
- 3) The dewar should hold a temperature of -185 degree celsius
- 4) System should have constructed with strong lightweight aluminum material.
- 5) System should have high thermal efficiency and ultra-low evaporative losses. Liquid nitrogen evaporation rate should be less than 0.2% per day.
- 6) System should support 1.2ml, 2ml and 5ml cryo vials on cane.
- 8) System should have 3 year vacuum warranty.
- 9) The warranty should be for 3 year.
- 10) System should have complied with Pressure Equipment Directive or equivalent.
- 11) System should be ISO/CE certified.

Specification for Liquid Nitrogen Dewar 10 liters

- 1) System should have capacity of 10 liters $\pm 2\%$.
- 2) System should be equipped with an electronic level gauge which gives an accurate reading of dewar contents.
- 3) System should have liquid nitrogen evaporation rate less than 0.8% per day.
- 4) System should be supplied with all the accessories required to function
- 5) System should have 3 year vacuum warranty.
- 6) System should have 3 year Warranty.
- 7) System should be ISO/CE certified.
- 8) Installation of Liquid nitrogen dewar should be a part of the offer

Specification for Liquid Nitrogen Dewar 25 liters

- 1) System should have capacity of 25 liters $\pm 5\%$ with wide neck opening for storing the samples. Diameter of neck opening should be 90mm $\pm 2\%$.
- 2) System should be compatible for the storage of Biological materials with excellent holding time.
- 3) The dewar should hold a temperature of -185 degree celsius
- 4) System should have constructed with strong lightweight aluminum material.
- 5) System should have Static hold time greater than 180 days.
- 6) System should have field working time greater than 110 days.
- 7) System should have high thermal efficiency and ultra-low evaporative losses. Liquid nitrogen evaporation rate should be less than 0.2% per day.
- 8) System should have internal canisters for the storage of straws and cryovials on canes.
- 10) System should support to store the 50ml falcon tubes on cane
- 10) System should support 1.2ml, 2ml and 5ml cryo vials on cane.



- 11) System should support 0.25cc and 0.5cc straw storage on cane.
- 12) System should be equipped with heavy duty roller base for easy mobility.
- 13) System should have heavy duty lockable enclosure.
- 14) System should have 5 year vacuum warranty.
- 15) The warranty should be for 3 year.
- 16) System should have complied with Pressure Equipment Directive or equivalent.
- 17) System should be ISO/CE certified.
- 18) Installation should be a part of the offer.

