

Ref: INS/L-5508/2019-2020(E)

Date : 04/10/2019

ENQUIRY

Dear Sirs,
Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		DELI REFRIGERATOR (Sliding Door) (Specification Attached)		3.00	No.

Remarks : TWO PART TENDER- The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.

Note :

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for _____" Ref No : _____. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 18/10/2019 till 02:00pm

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & Warranty period.

7. Delivery to be made to our Stores. Please indicate charges, if any extra. Transit Insurance should be done upto inStem Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, Please state so and then offer alternative to quality/Specifications.

9. Payment : within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs Duty (Except ad valorem duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.

11a. The Price quoted against this RFQ should be extended to Bangalore Life Science Cluster (BLiSc) of NCBS, InStem and C-CAMP for placing repeat order as per norms, by any one or all institutes of Bangalore Life Science Cluster (BLiSc)**inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification. **The Tenders to be quoted in foreign currencies & any other currencies approved/traded by RBI - USD/Euro/JPY/GBP/SGD/CAD/INR.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: **INS/L-5508/2019-2020(E)**

Date : 04/10/2019

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to read 'Elson Antony Paul'.

Elson Antony Paul

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



**2 PART TENDER FOR DELI REFRIGERATOR (Sliding Door) Technical Specifications - quantity 3 Nos.
Tender Reference INS/L-5508/2019-2020(E)**

SPECIFICATIONS FOR DELI REFRIGERATOR (Sliding Door)
1. Effective Capacity : 480 to 490 liters capacity
2. Exterior Cabinet : Galvanized steel with baked-on finish
3. Interior Cabinet : Stainless steel
4. Insulation : Rigid polyurethane foamed-in-place (HFC free)
5. Cooling Method : Forced cool air circulation
6. Compressor : Hermetic type, output: 200W
7. Evaporator : Fin tube type
8. Condenser : Wire and tube, Tube on sheet
9. Refrigerant : R-134a (HFC)
10. Fan motor : For cooling : output 3W X 1
11. Defrosting Method : Forced type fully automatic
12. Defrost heater : 73.8W x 2
13. Doors : Sliding Doors, Double layer glass with heat reflection film
14. Temperature range : 2°C to 14°C (ambient temperature - 5°C to 35°C no-load)
15. It should have Stable and reliable environment to control-the key to high-precision performance for exacting storage or research requirements. The refrigerator should be of good quality, previous user experinece will be considered.
16. It must be Microprocessor controlled
17. Refrigerator must provide a clear view of stored items through the large framed windows.
18. It must have double-sensor control system with alarm-exclusive sensor added to the temperature control sensor to always monitors the precise chamber temperature
19. It must have medical level safety standards
20. It must have a film to prevent shattered glass from scattering onto the floor
21. It must use HFC refrigerant and the foamed-in-place insulation must be HFC free.
22. It must be certified with ISO 9001 TUV SD, ISO 4001 JACO, UKAS MANAGEMENT SYSTEMS.
23. Warranty of 3 years. Installation by a Bangalore based engineer, who is available to rectify any instrument related issues within 2 working days of the notice. Details and contact of this engineer has to be provided with the tender.

**2 PART TENDER FOR DELI REFRIGERATOR (Sliding Door) Technical Specifications - quantity 3 Nos.
Tender Reference INS/L-5508/2019-2020(E)**

Terms and Conditions:

1. The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.
2. first sealed cover – Cover I, and super scribed as “Techno-commercial Bid” and should contain Complete Technical details of the Instrument offered (Specifications, Technical Parameters, Advantages, etc.,) .
3. The second sealed cover – Cover II super scribed 'Price Bid' should contain only rates (should be duly signed with seal and filled with date wherever necessary)
4. THESE TWO COVERS SHALL BE AGAIN PUT INTO A SINGLE WAX SEALED COVER super scribed “**DELI REFRIGERATOR (Sliding Door) -Qty 03 Nos.**” and should reach INSTEM on or before **18-10-2019** before **14:00 hrs**”. This should be addressed to the Purchase Officer, inStem, NCBS, GKVK Post, Bellary Road, Bangalore – 65.
5. The tender to be quoted in foreign currencies & any other currencies approved/traded by RBI-USD/Euro/JPY/GBP/SGD/CAD/INR.
6. If the items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.



**2 PART TENDER FOR DELI REFRIGERATOR (Sliding Door) Technical Specifications - quantity 3 Nos.
Tender Reference INS/L-5508/2019-2020(E)**

INFORMATION TO TENDERERS

The Tender shall be evaluated under 2 (Two) Bid System

- a. Technical Bid
- b. Financial Bid

SPECIFICATIONS FOR DELI REFRIGERATOR (Sliding Door)	Score
1. Effective Capacity : 480 to 490 liters capacity	15
2. Exterior Cabinet : Galvanized steel with baked-on finish	
3. Interior Cabinet : Stainless steel	
4. Insulation : Rigid polyurethane foamed-in-place (HFC free)	
5. Cooling Method : Forced cool air circulation	10
6. Compressor : Hermetic type, output: 200W	
7. Evaporator : Fin tube type	
8. Condenser : Wire and tube, Tube on sheet	
9. Refrigerant : R-134a (HFC)	25
10. Fan motor : For cooling : output 3W X 1	
11. Defrosting Method : Forced type fully automatic	
12. Defrost heater : 73.8W x 2	
13. Doors : Sliding Doors, Double layer glass with heat reflection film	20
14. Temperature range : 2°C to 14°C (ambient temperature - 5°C to 35°C no-load)	
15. It should have Stable and reliable environment to control-the key to high-precision performance for exacting storage or research requirements. The refrigerator should be of good quality, previous user experinece will be considered.	
16. It must be Microprocessor controlled	
17. Refrigerator must provide a clear view of stored items through the large framed windows.	10
18. It must have double-sensor control system with alarm-exclusive sensor added to the temperature control sensor to always monitors the precise chamber temperature	
19. It must have medical level safety standards	
20. It must have a film to prevent shattered glass from scattering onto the floor	
21. It must use HFC refrigerant and the foamed-in-place insulation must be HFC free.	20
22. It must be certified with ISO 9001 TUV SD, ISO 4001 JACO, UKAS MANAGEMENT SYSTEMS.	



Institute for Stem Cell Biology and Regenerative Medicine

2 PART TENDER FOR DELI REFRIGERATOR (Sliding Door) Technical Specifications - quantity 3 Nos. Tender Reference INS/L-5508/2019-2020(E)

23. Warranty of 3 years. Installation by a Bangalore based engineer, who is available to rectify any instrument related issues within 2 working days of the notice. Details and contact of this engineer has to be provided with the tender.	
	100
Evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening, will qualify for Price Bid opening. Thereafter, Financial proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.	

