



Ref: INS/L-5536/2017-2018(Y)

Date : 22/09/2017

### ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Nano Drop Spectrophotometer (As per specification attached)	-	1.00	No.

**Remarks :**

**Note :Quotation in USD, GBP, YEN, EURO, CND, SGD, ect will be acceptable.**

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for \_\_\_\_\_" Ref No : \_\_\_\_\_. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS  
11/10/2017 till 5.30 p.m.**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5.All duties,taxes,surcharge and cess as currently applicable must be sated in your quotation,seperately.Otherwise your quote is liable to be rejected.

6.Your quotation should indicate delivery period & Warranty period.

7.Delivery to be made to our Stores.Please indicate charges,if any extra.Transit Insurance should be done upto inStem Stores.

8.If you are unable to supply the quality,specifications or brand as mentioned in our enquiry,Please state so and then offer alternative to quality/Specifications.

9.Payment :within one month after delivery & acceptance/satisfactory installation.

10.Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11.Since we are a public funded research institution,we are exempted from paying Customs Duty(Except advolerum duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments,technical instruments,equipments(including computers),their accessories,spares,consumables and software.Hence,please offer your prices taking this option into consideration.



**Institute For Stem Cell Biology and Regenerative Medicine**  
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5536/2017-2018(Y)

Date : 22/09/2017

**12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.**

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell  
Biology and Regenerative Medicine

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in





Institute For Stem Cell Biology And Regenerative Medicine

Ref No: INS/L-5536/2017-2018(Y)

### Specifications of NANO DROP Spectrophotometer

The system should have the UV VIS spectral range of 190 -840 nm

The light source of the system should be Xenon flash lamp and should have long rated life.

The system should have pedestal micro volume measurement.

The system should be supplied with the necessary PC with the instrument

The system should have the Pre-programmed methods to measure absorbance for RNA, Proteins for quick start and other standard application for Biological sample analysis.

The system should have the Fast measurement time of less than 10 seconds

The system should have option for Blank,Re-Blank.

The system should have the wavelength accuracy of  $\pm 1$  nm

The system should have absorbance accuracy of  $\pm 5$ nm.

The system have absorbance range of 0.5 to 300.

The system should be able to measure the minimum sample volume 0.5 micro liter.

The system should be supplied with pedestal cleaning solutions.

The system should be supplied with calibration solutions.

The system should support Data Exporting option (Excel, notepad and PDF Formats, etc.)

The system should be supplied with all the accessories required to function

The system should have the warranty for 3 years

The system should be included with User-friendly software

The system should be CE/ISO Certified

