

2 PART TENDER FOR Cage, Rack & Bottle washer -Qty 01 No.

Tender Fee – Rs.700/-, EMD: Rs. 3,06,000/-
Cost of Tender: Approx. Rs. 153 Lakhs
Last date for Sale of Documents: 15-11-2019 till 14.00hrs
Last date for submission: 18-11-2019 till 15.00hrs
Due date for opening bids: 18-11-2019 at 15.30hrs

GENERAL CONDITIONS

(Tender documents consist of 16 pages, Pg. 1 to Pg.15 – Technical Bid, Pg. 16- Price Bid)

Important Instructions: The bids shall be enclosed in an envelope and sealed duly marked “Cage, Rack & Bottle washer -Qty 01 No.” Ref. No. INS/L-5557/2019-2020 (Y)”; addressed and to be mailed to “The Purchase Officer”. The bids are liable to be rejected if the sealed envelope is not addressed to “The Purchase Officer” with Tender Ref. No. and Item Description. Offers delivered in person shall be deposited in the Tender Box Labelled as 'TENDER BOX FOR PURCHASE TENDER' kept in the Ground Floor, Reception at Administration Building. If the bids are sent through courier or mail, it should reach by submission Date and Time and inStem will not responsible for the delay.

Important instruction for the Tenderers: -

All tenders shall be made in ENGLISH only.

The details in regard to technical specification and other terms and conditions should be cogent and clear to the extent possible.

Sealed tenders are invited at Institute for Stem Cell Biology and Regenerative Medicine -inStem, National Centre for Biological Sciences, GKVK Post, Bellary Road, Bangalore – 560 065 under **2 cover system** from reputed **OEM (Original Equipment Manufacturer) or reputed Manufacturers / Authorised Dealers for Supply, Installation, Commissioning and Technical Support for the following: -**


Cage, Rack & Bottle washer -Qty 01 No.

The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially **LOWEST BIDDER** shall be the first preferred Vendor for award of Order.

I) The EMD amount should be put in **first sealed cover – Cover I**, and super scribed as “**Techno-commercial Bid**” and should contain -

1. Acceptance of Technical specifications, **Annexure- A (page 9)** and terms and conditions. Tender document to be enclosed with Date, Signature and Seal in every Page.
2. Complete Technical details of the Instrument offered (Specifications, Technical Parameters, Advantages, etc.,)
3. Supplier profile & Schedule of Experience – **Annexure – B**
4. Supplier must describe in detail the technical support they will be able to provide in Bangalore. Only those companies will be considered who have engineers based in Bangalore who have been trained on the machines being quoted for, prior to the date of installation.
5. Suppliers must provide complete list of publications arising from use of their machines, in which the machines have been used for the applications listed. Enclose Data Sheet and Sample Analysis (if any)
6. Details of Warranty Services.
7. Details of “After Sales Service” and “Factory Trained Engineers” available in Bangalore.
8. Xerox copies of the Purchase Orders for having supplied similar Instruments in India.
9. Schedule of deviation from specifications / conditions – **Annexure C**.
10. Any other information the bidder like to provide with Date, Signature and Seal





Ref: INS/L-5557/2019-2020(Y)

Tender Notice No.018/2019-2020

**(Annexures should be duly signed and filled with date wherever necessary)
(Please attach additional sheet(s), wherever necessary)**

II) The **second sealed cover** – **Cover II** super scribed '**Price Bid**' (page 16) should contain only rates **(please attach additional sheet(s), wherever necessary)** (should be duly signed with seal and filled with date wherever necessary)

THE ABOVE MENTIONED TWO COVERS SHALL BE SEALED ON THE OUTSIDE WITH WAX SEAL BEARING THE LOGO/NAME OF THE COMPANY SUBMITTING THE BID.

THESE TWO COVERS SHALL BE AGAIN PUT INTO A SINGLE WAX SEALED COVER super scribed "**Cage, Rack & Bottle washer -Qty 01 No.**" and should reach **INSTEM on or before 18-11-2019 before 15.00 hrs**". This should be addressed to the Purchase Officer, inStem, NCBS, GKVK Post, Bellary Road, Bangalore – 65. The Techno – Commercial Bid will be opened on **18-11-2019 at 15.30 hrs**. On the date of tender opening (i.e. on **18-11-2019**), only the **Techno-Commercial Bids** shall be opened in the presence of attending tenderers.

The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% and above will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.

Earnest Money Deposit shall be submitted along with the "Techno-Commercial Bid" in the form of a Demand Draft drawn in favor of "Institute for Stem Cell Biology and Regenerative Medicine, payable at Bangalore" and the DD should be from a Commercial Bank/ Foreign Bank (First class Bank). Alternatively, a Bank Guarantee from a Commercial Bank/ Foreign Bank (First class Bank) may be provided (no other mode of payment will be accepted). The Bank Guarantee should be valid for 6 months from the date of opening. Bids not accompanied with Earnest Money Deposit shall be rejected. The EMD shall be refunded to the unsuccessful bidders once the order is released on the successful bidder. The Techno-Commercial bids shall be evaluated subsequently and only the shortlisted firms Price Bid will be opened.

1.1 Quotations must be submitted giving complete details using enclosed tender papers.

1.2 The rates quoted should remain valid for a period of 180 days from the date of Price Bid opening.

1.3 Each page of the tender except the Price & Delivery part shall be on printed letterheads or forms and bear the signature, date, name and designation of the person signing the offer. If they are not on letterheads, a rubber stamp indicating full name, address and phone No., Telex No., Fax No. etc. of the firm shall be affixed at the end of each page. The price & delivery part shall be as Annexure B attached.

1.4 This tender document is not transferable. Only the party to whom the tender documents have been issued shall be entitled to quote.

1.5 Bids containing erasures or alterations are liable to be rejected unless countersigned by the Authorised signatory.



- 1.6 All rates and total amount should be written both in figures and in words and if there is any discrepancy between the two, the lowest amount only will be considered.
- 1.7 Tenders will be opened in the presence of attending tenderers on **18-11-2019 at 15.30 hours**. In the event the due date is declared as holiday, the tender will be opened on the following working day.
- 1.8 If the item offered is to be imported, arrangements for import will be made by the Institute. Import duty and customs clearance will be under purchaser's scope.
- 1.9 We reserve the right to place order for part/reduced quantity than what is specified in the tender and also reserve the right to split the order to more than one supplier.
- 1.10 Any deviation/substitution in regard to the technical specification must be indicated in Annexure C of this tender document. Otherwise it shall be binding on the bidder to supply the items as specified in this tender specification.
- 1.11 All bids are to be submitted before the due date and time. **The bids shall be enclosed in an envelope and sealed duly marked "Cage, Rack & Bottle washer -Qty 01 No."; Ref. No. INS/L-5557/2019-2020 (Y); and mailed to the Purchase Officer.** Offers delivered in person shall be deposited in the Tender Box kept in the office. **Due date for opening the bids: 18-11-2019 (15.00 hours).**
- 1.12 Please return the tender papers including Conditions of Tender as well as the Annexures with your signature, rubber stamp and date affixed on each page.
- 1.13 All bids in response to this invitation of tender should be submitted in a manner and method specified above. Tender which do not comply with the above conditions are liable to be rejected.
- 1.14 Late and delayed tenders will not be considered. Therefore, tenderers shall ensure that the tender reaches the Purchaser on or before the due date and time stipulated for receipt of bids.
- TENDERS RECEIVED LATE OR AFTER THE DUE DATE WILL NOT BE CONSIDERED. INSTEM RESERVES THE RIGHT TO ACCEPT, REJECT ANY OR ALL TENDERS WITHOUT ASSIGNING ANY REASONS THERE OF.**
- 1.15 Individuals signing the bid form and other supporting documents must specify the capacity in which they sign, like -
- Whether signing as a Sole Proprietor of the firm or his attorney.
 - Whether signing as a partner of the firm or his attorney
 - Whether signing for the firm as Agent.
 - Whether signing as Director of a Limited Company.

2. CATALOGUE/TECHNICAL LITERATURE

All necessary catalogue/drawing literature/data and details of item/s as are considered to be essential for full and correct evaluation of the bid shall invariably accompany the bid.

3. BID GUARANTEE / EARNEST MONEY DEPOSIT: -

Bid Guarantee amount details are as below: -

Sl. No.	Item Description	BGA/EMD Amount
1.	Cage, Rack & Bottle washer -Qty 01 No.	INR 3,06,000/-

The EMD/BGA shall be submitted by a DD (for outstation firms) or Banker's cheque (in case of local firm) from a Commercial Bank/ Foreign Bank (First class Bank) along with the bid, drawn in favor of "Institute for Stem Cell Biology and Regenerative Medicine, Bangalore". Alternatively, the EMD amount may be submitted by way of Bank Guarantee from a Commercial Bank/ Foreign Bank (First class Bank) valid for 6 months (no other mode of payment will be accepted). The BGA amount will be forfeited if the successful bidder fails to accept the Letter of Intent/ Purchase order or withdraws or amends, impairs or derogates from the tender in any respect within the period of validity of this tender. The offers are liable to be rejected, at the discretion of the Centre, if they are not accompanied with BGA. No interest shall be payable by INSTEM for BGA amount. The BGA shall be refunded to the unsuccessful bidders once the order is released on the successful bidder.

4. AUTHORITY OF AGENTS

In case where a bid is submitted by an Indian Agent on behalf of his foreign principal, such bid should be supported with a letter of Authority from the principal that the Indian Agent has been authorised to submit the bid on behalf of the principal. The indigenous manufacturers shall submit their bids directly.

5. PRICE

The tender to be quoted in foreign currencies & any other currencies approved/traded by RBI-USD/Euro/JPY/GBP/SGD/CAD/INR. The price/s quoted shall be firm till the complete execution of the order. All details relating to price, price breakup, inland transportation, documentation, taxes and duties, levies, Road/AIR/Marine freight charges, delivery terms (ex-works/F.O.R/F.O.B/CIP) mode of payment, mode of Dispatch, Insurance, Agency Commission, if any, should be clearly stated. For indicating the price, the tenderers may choose any/all of the following:

- Ex-works (all other charges to be indicated separately).
- F.O.R. site (i.e. Freight, Packing & Forwarding, loading on to the transport, documentation etc. included.) Internal transportation, Freight, Insurance, etc. to be shown separately.
- F.O.B (cost of goods, Packing & Forwarding, Inland Transportation, Documentation, etc. till the item is loaded on to the cargo carrier). Freight & Insurance charges to be separately indicated.
- CIP (cost of goods, packing & forwarding, documentation, freight, insurance, etc. all included). However, freight & insurance charges to be indicated separately.

INSURANCE TO BE COVERED TILL inStem STORES.

- The taxes (State, Central, Turnover tax, Works Contract Tax, etc.). Please specify which are applicable. The duties and other levies, freight, insurance shall be stated clearly and separately. Also please mention whether the same is included in the price/s quoted.

7. VALIDITY OF BIDS

The bids should be valid for a period of 180 days from the date of opening of the Price bids. Bids with shorter validity period are liable for rejection.



8. DELIVERY

The tenderer should clearly mention the time required for supplying the item. The period of delivery will be counted (a) from the date of receipt of the order in case of Indian supplier and (b) from the date of opening of letter of credit/receipt of order (in the case of sight draft/advance draft) for foreign suppliers. The delivery date is the date at which the equipment should be delivered at INSTEM.

9. PACKING

The item should be packed appropriately so that it can sustain transit hazards, multiple landing, warehousing, etc. during transit.

10. PLACE OF DELIVERY

In the case of items indigenously manufactured, ALL DELIVERIES shall be effected to the Institute for Stem Cell Biology and Regenerative Medicine, National Centre for Biological Sciences, GKVK, Bellary Road, Bangalore-560065, whereas the foreign suppliers shall effect delivery up to Bangalore Airport, India by AIR only unless otherwise specified.

11. GUARANTEE/WARRANTY

The item/s covered under this tender shall be subject to a guarantee for trouble free performance, workmanship, material etc., fulfilling the specifications mentioned in this tender for min. **36 months** from the date of commissioning. If any defect is found in the material, workmanship or performance during the guarantee period the same may either be repaired/replaced by the supplier as the case may be free of charge. The guarantee period for replacement of parts or repair work shall be same as above. A guarantee certificate to this effect should be forwarded to us with your invoice. For order placed on foreign suppliers, the supplier shall specifically confirm that their Indian representatives, if any, will provide with after sales service and will attend to any repairs or technical problems that may arise.

12. PERFORMANCE B/G

The successful bidder shall have to execute a performance B/G in accordance with the guarantee/warranty for 10% value of the order. This B/G has to be executed on an appropriate value of stamp paper in terms of a bank guarantee drawn on any Nationalized Bank or first class Foreign Bank and shall remain valid till the completion of the Defect Liability period/warranty period, with **6 months** claim period.

13. TAXES AND DUTIES

The Centre shall deduct all taxes and duties, as applicable, from time to time from the bills payable.

14. PAYMENT TERMS

As per our Centre, 90% of the payment shall be made through Sight Draft (documents through bank) and balance 10% payable after installation, trial run and subject to acceptance by the group head through SAI and all other relevant documents, with Performance Bank Guarantee.

All claims shall cease as per the Limitation Act.

15. CLARIFICATIONS

After opening the bids, if it becomes necessary for the purchaser to seek clarifications from the bidders, the same will be sought from the bidders. In such an event, the bidders will furnish all technical information / clarifications to the purchaser to reach on or before the due date fixed

for that purpose, indicating the Purchaser's tender reference. If the technical clarifications sought do not reach on or before the date fixed, the bids shall be summarily rejected without any further notice.

16. **RISK CLAUSE**

Notwithstanding the other terms therein, the Centre at its option will be entitled to terminate the contract and to avail from elsewhere; at the risk and cost of contractor; either the whole of the contract or any part which the contractor has failed to perform within the time stipulated or if the same performance is not available, the best and the nearest available substitute thereof. The contractor shall be liable for any loss which the Centre may sustain by reason of such risk contract in addition to penalty.

17. **DISPUTE AND RESOLUTION**

Any dispute or differences between the parties that cannot be settled by mutual discussion at appropriate levels shall be referred to the sole arbitration of the Director, INSTEM or his nominee and his decision in the matter shall be final and binding upon the parties to the dispute. The venue of arbitration proceedings shall be Bangalore. In respect of any matters pertaining to such arbitration, the courts of law in Bangalore will have exclusive jurisdiction.

18. **OTHER TERMS**

a) If electrical/ AC/ technical works contract, appropriate license defining the required expertise from the approved Licensing authority.

b) The Centre, will at its discretion ask for a solvency certificate from your bankers.

19. **DEFINITION AND MEANINGS**

In constructing these conditions, specifications, etc. in the tender document or the Annexures/Appendices the following words shall mean herein assigned definitions except where the subject context is otherwise stated.

PURCHASER: Shall mean the Institute for Stem Cell Biology and Regenerative Medicine


BID/TENDER: Shall mean the proposal/document that the BIDDER submits in the requested and specified form or otherwise along with Annexures, Appendices, etc.

BIDDER/TENDERER: Shall mean the firm/party who quotes against an enquiry.

CONTRACTOR/SUPPLIER: Shall mean the party to whom a Work Order/Purchase Order is awarded to undertake all or a part of the work covered by this tender document as well as an amendment orders relating to this tender issued by the Purchaser and shall include his/their legal representative, assignee/s or successor/s.

CONTRACT: Shall mean and include the articles of agreement, Declaration form, the general and special conditions, the





Ref: INS/L-5557/2019-2020(Y)

Tender Notice No.018/2019-2020

Annexures, the Schedule of Quantities & rates and the specifications attached hereto and the drawings, if any.

ORDER VALUE:

Shall mean total value of the Purchase Order/Work order issued against this tender item including taxes, levies, etc.

20. **TECHNICAL SPECIFICATIONS**

See Annexure – 'A'

21. For imported items, final price after totaling the prices of the individual items has to be mentioned with estimated freight, insurance, agency commission charges, etc. The price shall be FOR CIF BANGALORE AIRPORT.
22. INSTEM reserves the right to split the quantities or reject one or more offers in full or part without any reasons. Therefore, INSTEM decision is final and binding.
23. The quantity mentioned in the tender may be decreased/increased while ordering. However, the price quoted should be firm irrespective of change in the quantity.
24. **ACCEPTANCE OF TENDERS**
INSTEM does not pledge itself to accept the lowest/ any tenders and reserves to itself the right to accept the whole or part of the tenders or a part of the quantity offered.
25. The firm may quote in Foreign Currency for direct import on CIF basis as well as in Rupees on FOR INSTEM basis for local supply. Please indicate insurance, freight and other charges separately for direct import.
26. Your Service Engineer should be fully trained to install the equipment and capable of maintaining the equipment during / after the warranty period.


27. **LIQUIDATED DAMAGES:**

The successful bidder will have to agree that in case the item is not supplied within the agreed delivery schedule and after a grace period of seven days, then Liquidated Damages (not in terms of penalty) will be imposed automatically and be deducted from their bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.

28. **Micro and Small Enterprises (MSEs):**

- a. Micro and Small Enterprises (MSE) must, along with their offer, provide proof of their being registered as MSE (indicating the terminal validity date of their registration) for the item tendered, with any agency mentioned in the notification of the Ministry of Micro, Small and Medium Enterprises (Ministry of MSME)
- b. The MSEs are exempted from payment of earnest money and tender fees subject to furnishing of relevant valid certificate for claiming exemption as per privilege rules of Government of India.





Ref: INS/L-5557/2019-2020(Y)

Tender Notice No.018/2019-2020

c. The bidder submits registration of Udyog Adhar Memorandum (UAM) by Ministry of Micro Small and Medium Enterprises (MSME) vendors on Central Public Procurement Portal (CPPP). The bidders who fail to submit UAM number shall not be able to avail the benefits available to MSEs as contained in Public Procurement Policy for MSEs Order 2012 issued by MSME.

Since INSTEM is a Public Funded Research Organisation, registered under Government of India Notification No.51/96 – Customs and 10/97 – for Central Excise is applicable for the items listed in the notification.

For and on behalf of
Institute for Stem Cell Biology and Regenerative Medicine




Purchase Officer

GKVK Post, Bellary Road, Bangalore 560 065, India
Phone +91-80-23666343 /344/345/346 . Telefax +91-80-23636662
purchase@ncbs.res.in . www.ncbs.res.in



ANNEXURE-A-SPECIFICATIONS
Cage, Rack & Bottle washer -Qty 01 No.
TECHNICAL SPECIFICATIONS:

Supply and installation of electrically operated heavy duty, large capacity and hydro-spray washer with the capability of washing most of the Equipments and materials used in the lab animal facility such as animal cages, lids, trolleys, conventional racks, IVC racks, water bottles and other miscellaneous items used in the care of laboratory animals (mice, rats, guinea pigs and rabbits etc.). The Same machine should have the capability to decontaminate by using H2O2 to all heat sensitive instruments, like IVC AHU, Cage Changing Station, Bio Safety Cabinets etc.
The whole machine should have exterior stainless covering along with applicable thermal insulation. The washing chamber, technical compartment, panels, spray headers or oscillating arms, wash and rinse tank, piping, door panels should be made of at least AISI 304 stainless steel. Pump housing and impeller, Process valves should be made of 316 stainless steel.
1. Washing Chamber: The washing chamber should have minimum dimensions to accommodate the racks of larger size and other large Equipments easily for washing/cleaning. The floor should be made of small stainless steel and properly designed to act as a primary debris filter and it should facilitate easy cleaning and maintenance. The washing cycle should have minimum 4 phases like, washing with temperature 50-60 degree, dripping, Rinsing with 82 degree and steam extraction
2. Doors: This machine should have double doors and should work as pass through system between clean and dirt area. Doors should be with double-layer tempered glass and must consist of two 6mm layers with Poly-vinyl-butryal film in-between. Doors (frames, glass and gaskets) to provide 100% airtight application, full vision inside the chamber, and reduced noise and heat transmission when the machine is running. The door opening on clean and dirty side needs to be provided as per space availability. Doors should facilitate the easy handling and management of daily operations smoothly. The appropriate provisions should be present to assure the operator safety during operations. It should facilitate unloading of the material in to clean side or dirty side again as per the procedural requirements at site. Door should come fitted with the inflatable gasket to ensure the air-tightness of the chamber during operations. During a standard cycle the maximum temperature measured on the external glass cannot exceed 42°C.
3. Mechanical Core or Technical Compartment: The side of the mechanical core or technical compartment towards left or right side needs to be provided as per the space availability.
4. Monitoring, Control and Data Management System: The machine should have the built in monitoring system that automatically controls all required process operations and functions. The control system or operator interface should be in place to program and pre-set several different cycles as per various washing requirements. It should be user friendly and allows easy parameters setting/controlling. It should have programmable security access. The machine should have the appropriate data management system in place.
5. Spraying System: The machine should come with oscillating arms fitted all along the side walls of the chamber, with wash and rinse nozzles and the rotation of the arms should cover all the wash surfaces. The WASH & RINSE should come with a separate circuit to avoid the cross contamination.
6. Water Tanks: Separate wash and rinse tank. Include top inspection latch. Equipped with level control, automatic fill, overflow safety and temperature control. Wash water tanks with radius corners and bottom slope to outlet. All tanks to be insulated and sealed.
7. Water Pumps: Wash pump should deliver water from wash tank to washing circuit of jet-spray manifold with min. pressure at wash nozzles 1.5 - 2bar.
The Rinse pump should deliver fresh clean and pressurized water to rinse circuit of jet-spray manifold. Used rinse water to be returned to wash tank for re-use and conservation of heat load. Min. pressure at rinse nozzles should be 3-4 bar.
Recalculating pump should deliver water from rinse tank at base of wash chamber back to wash tank
8. Self-Cleaning Filter: Tool-free self-cleaning wash water filtering system should be available and back-flushed at the end of each washing phase to ensure the debris removal from the mesh during the self-flushing phase. No any filters equipped in the floor's grid in order to avoid deposition of bedding inside the chamber
9. Detergent Dosing System: The machine should be equipped with a detergent dosing system with independent pumps for acid and alkaline washes.
10. Bottle Washing System: The machine should come with a dedicated set-up for bottle washing and allows washing of the bottles crates on the required presentation rack/carts.
11. Decontamination System: The machine should come with a dedicated on-board VHP Generator with Fast Aeration Kit and Catalyst Unit. Integral control system and software to interface between generator and system PLC. The Chamber should have the provision for internal power socket. The system able to reduce the cycle to One Hour. Portable H2O2 sensor should be supplied. The Rack washer chamber should ensure the safety of the decontamination cycle.



Ref: INS/L-5557/2019-2020(Y)

Tender Notice No.018/2019-2020

12. Bid should include all the required accessories required for the machine to maximize the wash load like presentation racks (dedicated/universal for cages, lids etc.), bottle processing racks/carts, baskets/crates and other accessories as applicable.
13. Safety System: The machine should have the personal safety system in place such as emergency bars/stop cables inside the chamber and emergency switches outside the chamber in both clean & dirty sides as an integrated option.
14. Noise level generated by the machine should be less than 80 decibel (dB)
15. Electricity requirements should be Approx. 70KW and the washing cycle should consume less than 50 Liters water per cycle
16. Please provide the details about the Wash cycles validation support and accomplishment.
17. The required wash cycle should be completed in less than 5 minutes and also please specify the different cycle times as applicable for this machine.
18. Please mention or provide the applicable international standard certifications for quality, safety and environmental impact for this machine.
19. At least the following documentation should be included along with the machine:
a. User and Maintenance Manual
b. P&I Diagram
c. Pneumatic diagram
d. Wiring Diagram
20. The prospect vendor should specify and provide full validation documents (IQ, OQ, PQ etc.) as applicable.
21. Pre-installation and utility requirements for installation and running the system should be clearly mentioned and considered while submitting the bids. There is no provision of centralized utilities such as steam, hot water and compressed air at site. Please mention the specific utility systems (if any) required for running this machine to get better output. Please mention the brand or company name with capacity and also the budgetary quote needs to be provided for reference for all those items or systems required as utility.
22. Please indicate the year in which the model was introduced in the market and confirm whether the spares and consumables for the machine would be available for a minimum period of 10 years.
23. Detailed literatures with technical specifications and features for the machine and all the related accessories to be enclosed with the offer.
24. Prospective bidders are requested to visit the site and offer suitable bids in the required format along with literatures.
25. Manufacturer should have standard certifications (ISO 9001 and ISO 14001).
26. The machine should have the applicable International standards for quality, safety and environmental impact such as AKKAB Guidelines. No any self-declaration accepted, only certificate from a third party company
27. Please ensure that the specifications mentioned in the offer must cover all the parameters listed in our enquiry and should attach the supporting documents. Unspecified parameter will be treated as non-compliant.
28. Installation testing, checking of specifications and the validation should be done at the specified area. Training of users to be done free of cost at the site.
29. Please indicate the year in which the model was introduced in the market and confirm whether the spares and consumables for the system would be available for a minimum period of 10 yrs. Warranty on overall machine and spares for 3 years.
30. Pre-installation and utility requirements for installation and running the system should be clearly mentioned.
31. Availability of local service support and response time for a service call during and after the warranty to be specified.
32. List of users in India and abroad of the similar models (Cage & Rack Washers) as the one(s) offered along with the names, addresses, telephone numbers and mail ID's to be enclosed separately. There should be minimum 5 reference up & running installations in India in LAS applications, in that three should be from Govt. institution.
33. Technical presentation and Demo on the systems offered is to be made on request from the Institute. In case any of the invited parties fails to demonstrate their compliance as per the demanded specifications, the party will be treated as disqualified.



INFORMATION TO TENDERERS

The Tender shall be evaluated under 2 (Two) Bid System

I Technical Bid


II Financial Bid

TECHNICAL SPECIFICATIONS & EVALUATION CRITERIA WITH MARKS FOR 2 PART TENDER FOR “Cage, Rack & Bottle washer -Qty 01 No”.

TECHNICAL EVALUATION CRITERIA WITH MARKS	
Technical Requirement	Marks
Supply and installation of electrically operated heavy duty, large capacity and hydro-spray washer with the capability of washing most of the Equipments and materials used in the lab animal facility such as animal cages, lids, trolleys, conventional racks, IVC racks, water bottles and other miscellaneous items used in the care of laboratory animals (mice, rats, guinea pigs and rabbits etc.). The Same machine should have the capability to decontaminate by using H2O2 to all heat sensitive instruments, like IVC AHU, Cage Changing Station, Bio Safety Cabinets etc.	25
The whole machine should have exterior stainless covering along with applicable thermal insulation. The washing chamber, technical compartment, panels, spray headers or oscillating arms, wash and rinse tank, piping, door panels should be made of at least AISI 304 stainless steel. Pump housing and impeller, Process valves should be made of 316 stainless steel.	
1. Washing Chamber: The washing chamber should have minimum dimensions to accommodate the racks of larger size and other large Equipments easily for washing/cleaning. The floor should be made of small stainless steel and properly designed to act as a primary debris filter and it should facilitate easy cleaning and maintenance. The washing cycle should have minimum 4 phases like, washing with temperature 50-60 degree, dripping, Rinsing with 82 degree and steam extraction	
2. Doors: This machine should have double doors and should work as pass through system between clean and dirt area. Doors should be with double-layer tempered glass and must consist of two 6mm layers with Poly-vinyl-butyl film in-between. Doors (frames, glass and gaskets) to provide 100% airtight application, full vision inside the chamber, and reduced noise and heat transmission when the machine is running. The door opening on clean and dirty side needs to be provided as per space availability. Doors should facilitate the easy handling and management of daily operations smoothly. The appropriate provisions should be present to assure the operator safety during operations. It should facilitate unloading of the material in to clean side or dirty side again as per the procedural requirements at site. Door should come fitted with the inflatable gasket to ensure the air-tightness of the chamber during operations. During a standard cycle the maximum temperature measured on the external glass cannot exceed 42°C.	
3. Mechanical Core or Technical Compartment: The side of the mechanical core or technical compartment towards left or right side needs to be provided as per the space availability.	
4. Monitoring, Control and Data Management System: The machine should have the built in monitoring system that automatically controls all required process operations and functions. The control system or operator interface should be in place to program and pre-set several different cycles as per various washing requirements. It should be user friendly and allows easy parameters setting/controlling. It should have programmable security access. The machine should have the appropriate data management system in place.	
5. Spraying System: The machine should come with oscillating arms fitted all along the side walls of the chamber, with wash and rinse nozzles and the rotation of the arms should cover all the wash surfaces. The WASH & RINSE should come with a separate circuit to avoid the cross contamination.	
6. Water Tanks: Separate wash and rinse tank. Include top inspection latch. Equipped with level control, automatic fill, overflow safety and temperature control. Wash water tanks with radius corners and bottom slope to outlet. All tanks to be insulated and sealed.	25
7. Water Pumps: Wash pump should deliver water from wash tank to washing circuit of jet-spray manifold with min. pressure at wash nozzles 1.5 - 2bar.	
The Rinse pump should deliver fresh clean and pressurized water to rinse circuit of jet-spray manifold. Used rinse water to be returned to wash tank for re-use and conservation of heat load. Min. pressure at rinse nozzles should be 3-4 bar.	
Recalculating pump should deliver water from rinse tank at base of wash chamber back to wash tank	



8. Self-Cleaning Filter: Tool-free self-cleaning wash water filtering system should be available and back-flushed at the end of each washing phase to ensure the debris removal from the mesh during the self-flushing phase. No any filters equipped in the floor's grid in order to avoid deposition of bedding inside the chamber	
9. Detergent Dosing System: The machine should be equipped with a detergent dosing system with independent pumps for acid and alkaline washes.	
10. Bottle Washing System: The machine should come with a dedicated set-up for bottle washing and allows washing of the bottles crates on the required presentation rack/carts.	
11. Decontamination System: The machine should come with a dedicated on-board VHP Generator with Fast Aeration Kit and Catalyst Unit. Integral control system and software to interface between generator and system PLC. The Chamber should have the provision for internal power socket. The system able to reduce the cycle to One Hour. Portable H2O2 sensor should be supplied. The Rack washer chamber should ensure the safety of the decontamination cycle.	
12. Bid should include all the required accessories required for the machine to maximize the wash load like presentation racks (dedicated/universal for cages, lids etc.), bottle processing racks/carts, baskets/crates and other accessories as applicable.	
13. Safety System: The machine should have the personal safety system in place such as emergency bars/stop cables inside the chamber and emergency switches outside the chamber in both clean & dirty sides as an integrated option.	
14. Noise level generated by the machine should be less than 80 decibel (dB)	
15. Electricity requirements should be Approx. 70KW and the washing cycle should consume less than 50 Liters water per cycle	25
16. Please provide the details about the Wash cycles validation support and accomplishment.	
17. The required wash cycle should be completed in less than 5 minutes and also please specify the different cycle times as applicable for this machine.	
18. Please mention or provide the applicable international standard certifications for quality, safety and environmental impact for this machine.	
19. At least the following documentation should be included along with the machine:	
a. User and Maintenance Manual	
b. P&I Diagram	
c. Pneumatic diagram	
d. Wiring Diagram	
20. The prospect vendor should specify and provide full validation documents (IQ, OQ, PQ etc.) as applicable.	
21. Pre-installation and utility requirements for installation and running the system should be clearly mentioned and considered while submitting the bids. There is no provision of centralized utilities such as steam, hot water and compressed air at site. Please mention the specific utility systems (if any) required for running this machine to get better output. Please mention the brand or company name with capacity and also the budgetary quote needs to be provided for reference for all those items or systems required as utility.	
22. Please indicate the year in which the model was introduced in the market and confirm whether the spares and consumables for the machine would be available for a minimum period of 10 years.	25
23. Detailed literatures with technical specifications and features for the machine and all the related accessories to be enclosed with the offer.	
24. Prospective bidders are requested to visit the site and offer suitable bids in the required format along with literatures.	
25. Manufacturer should have standard certifications (ISO 9001 and ISO 14001).	
26. The machine should have the applicable International standards for quality, safety and environmental impact such as AKKAB Guidelines. No any self-declaration accepted, only certificate from a third party company	
27. Please ensure that the specifications mentioned in the offer must cover all the parameters listed in our enquiry and should attach the supporting documents. Unspecified parameter will be treated as non-compliant.	
28. Installation testing, checking of specifications and the validation should be done at the specified area. Training of users to be done free of cost at the site.	
29. Please indicate the year in which the model was introduced in the market and confirm whether the spares and consumables for the system would be available for a minimum period of 10 yrs. Warranty on overall machine and spares for 3 years.	






Ref: INS/L-5557/2019-2020(Y)

Tender Notice No.018/2019-2020

30. Pre-installation and utility requirements for installation and running the system should be clearly mentioned.	
31. Availability of local service support and response time for a service call during and after the warranty to be specified.	
32. List of users in India and abroad of the similar models (Cage & Rack Washers) as the one(s) offered along with the names, addresses, telephone numbers and mail ID's to be enclosed separately. There should be minimum 5 reference up & running installations in India in LAS applications, in that three should be from Govt. institution.	
33. Technical presentation and Demo on the systems offered is to be made on request from the Institute. In case any of the invited parties fails to demonstrate their compliance as per the demanded specifications, the party will be treated as disqualified.	
Total	100
Evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.	





Tender Notice No.018/2019-2020

Please furnish list of major supplies/works undertaken in relevant to this tender during the past two years.



ANNEXURE - C


SCHEDULE OF DEVIATION FROM SPECIFICATIONS/CONDITIONS

All deviations from the specifications/conditions shall be filled in by the bidder in this schedule.

--

The bidder hereby certifies that the above mentioned are the only deviations from Technical Specification of this tender. [State NIL if no deviation is envisaged].

Signature	
Name	
Designation	
Name of the company	
Date	
Seal of the company	



Ref: INS/L-5557/2019-2020(Y)

Tender Notice No.018/2019-2020

COVER – II PRICE BID

Cage, Rack & Bottle washer -Qty 01 No.

Sl. No.	Item Description	Qty	Unit Price	Total Price

Please enter Unit Price, Total Price and Grand Total in figures as well as words. Unit price and Total Price may be entered below each item.

