

Ref: INS/L-5559/2019-2020(Y)

Date : 19/10/2019

## ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Isoflurane anaesthesia machine for Rodents (Specifications Attached)		2.00	Nos.

**Remarks : 2 PART TENDER:** The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.

1. The bids shall be enclosed in an envelope, and due date sealed duly marked "Tender for \_\_\_\_\_" Ref No : \_\_\_\_\_. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS  
04/11/2019 till 5.30 p.m.**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & Warranty period.

7. Delivery to be made to our Stores. Please indicate charges, if any extra. Transit Insurance should be done upto inStem Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, Please state so and then offer alternative to quality/Specifications.

9. Payment : within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs Duty (Except ad valorem duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.



**Institute For Stem Cell Biology and Regenerative Medicine**

Autonomous institute of the Department of Biotechnology, Government Of India



Ref: **INS/L-5559/2019-2020(Y)**


**Date : 19/10/2019**

**12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.**

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell  
Biology and Regenerative Medicine

  
Yesu R



GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666341/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in

**2 PART TENDER FOR Isoflurane anaesthesia machine for Rodents-Qty 02 Nos.**

**P.O. No: INS/L-5559/2019-2020(Y)**

**Tender Specification for Isoflurane anaesthesia machine for Rodents-Qty 02 Nos.**

<b>System should Include:</b>
Table top stand (13" H x 12" W x 10" D)
One inlet flowmeter, 0 to 1000 mL
One set of inlet/outlet caps
Charcoal filter canister
High-pressure oxygen hose with DISS fittings
Anaesthesia System with isoflurane vaporizer must include a 3-year warranty
Vaporizer should be for Isoflurane anaesthetic and cap filling type with level indicator. Suitable for Mice and Rats (including neonate mice)
It should be non-rebreathing circuit for single animal at a time
<b>Features should include</b>
Brand new vaporizer canisters
Never reconditioned or remanufactured
Durable stand, accurate, high-quality flowmeters, small footprint
Complete systems
Including all the tubing and connectors needed
Y-adapter tubing for easy transfer from induction chamber to anaesthesia mask
System must include suitable gas scavenging system
Y-splitter circuits for easy switch over between induction chamber and nose cone circuit
Should include suitable oxygen flow meter
Oxygen flush for purging induction chamber before opening it
<b>Add-ons</b>
Induction chambers and anaesthesia masks/ nose cones must be included (suitable sizes for mice and rats both), cost-effective Anaesthesia Starter Kit (for mice) that includes all the components necessary for our setup.
<b>3-years warranty</b> must be included. Vendor must have local service Center for timely servicing and calibration.
Technical presentation and Demo on the systems offered is to be made on request from the Institute. In case any of the invited parties fails to demonstrate their compliance as per the demanded specifications, the party will be treated as disqualified



**Terms and Conditions:**

1. The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.
2. **first sealed cover – Cover I**, and super scribed as **“Techno-commercial Bid”** and should contain Complete Technical details of the Instrument offered (Specifications, Technical Parameters, Advantages, etc.,)
3. The **second sealed cover – Cover II** super scribed **'Price Bid'** should contain **only rates** (should be duly signed with seal and filled with date wherever necessary)
4. THESE TWO COVERS SHALL BE AGAIN PUT INTO A SINGLE WAX SEALED COVER super scribed **“Isoflurane anaesthesia machine for Rodents-Qty 02 Nos.”** and should reach **INSTEM on or before 04-11-2019 before 5.30 P.M.** This should be addressed to the Purchase Officer, inStem, NCBS, GKVK Post, Bellary Road, Bangalore – 65.
5. The tender to be quoted in foreign currencies & any other currencies approved/traded by RBI-USD/Euro/JPY/GBP/SGD/CAD/INR.
6. If the items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.
7. **Original Equipment Manufacturer or Authorised Dealers should provide Authorized Distribution Certificate along with the Quotation.**



**INFORMATION TO TENDERERS**

**The Tender shall be evaluated under 2 (Two) Bid System**

**I Technical Bid**

**II Financial Bid**

**TECHNICAL SPECIFICATIONS & EVALUATION CRITERIA WITH MARKS FOR 2 PART TENDER FOR "Isoflurane anaesthesia machine for Rodents-Qty 02 Nos.**

<b>Specifications: Isoflurane anaesthesia machine for Rodents</b>	<b>Score</b>
<b>System should Include:</b>	<b>35</b>
Table top stand (13" H x 12" W x 10" D)	
One inlet flowmeter, 0 to 1000 mL	
One set of inlet/outlet caps	
Charcoal filter canister	
High-pressure oxygen hose with DISS fittings	
Anaesthesia System with isoflurane vaporizer must include a 3-year warranty	
Vaporizer should be for Isoflurane anaesthetic and cap filling type with level indicator. Suitable for Mice and Rats (including neonate mice)	
It should be non-rebreathing circuit for single animal at a time	<b>30</b>
<b>Features should include</b>	
Brand new vaporizer canisters	
Never reconditioned or remanufactured	
Durable stand, accurate, high-quality flowmeters, small footprint	
Complete systems	
Including all the tubing and connectors needed	
Y-adapter tubing for easy transfer from induction chamber to anaesthesia mask	
System must include suitable gas scavenging system	<b>35</b>
Y-splitter circuits for easy switch over between induction chamber and nose cone circuit	
Should include suitable oxygen flow meter	
Oxygen flush for purging induction chamber before opening it	
<b>Add-ons</b>	<b>35</b>
Induction chambers and anaesthesia masks/ nose cones must be included (suitable sizes for mice and rats both), cost-effective Anaesthesia Starter Kit (for mice) that includes all the components necessary for our setup.	
<b>3-years warranty must be included.</b> Vendor must have local service Center for timely servicing and calibration.	
Technical presentation and Demo on the systems offered is to be made on request from the Institute. In case any of the invited parties fails to demonstrate their compliance as per the demanded specifications, the party will be treated as disqualified	<b>100</b>
<b>Total</b>	
<b>Evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.</b>	



