

Ref: INS/L-5613/2020-2021(Y)

Date : 27/01/2021

## ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Biosafety cabinet (Specifications Attached)		1.00	No.

**Note: 2 PART TENDER:** The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for \_\_\_\_\_" Ref No : \_\_\_\_\_. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS  
11/02/2021**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & Warranty period.

7. Delivery to be made to our Stores. Please indicate charges, if any extra. Transit Insurance should be done upto inStem Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/Specifications.

9. Payment Term: 97% against supply; balance 3% on Installation subject to receipt of PBG for 3% Order Value.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs Duty (Except ad valorem duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.



Ref: INS/L-5613/2020-2021(Y)


Date : 27/01/2021

**12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.**

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Institute For Stem Cell  
Science and Regenerative Medicine



Yesu R



GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666341/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



**2 PART TENDER FOR Biosafety Cabinet- Qty 01 No.**

Ref No: INS/L-5613/2020-2021(Y)

**Tender Specifications:**

<b>Biosafety Cabinet - Technical Specifications</b>
The Bio safety cabinet should be of high quality, imported, Type A2 in which 70% Air should be re-circulated and 30% of the air should be exhausted
The Bio Safety Cabinet must include two DC motors. High power consuming /high heat output AC motors must not be used
The motor must automatically adjust the airflow speed without the use of a damper to ensure continuous safe working conditions, even without maintenance adjustment , inflow and outflow should be independent control
In order to preserve safety to the user and the environment, the exhaust blower on the cabinet must continue operating when the supply blower stops working. If the exhaust blower should fail, the supply filter should also be turned off.
In order to ensure consistent and reliable down flow velocity across the supply HEPA filter over the life of the cabinet, the cabinet must use a pressure sensor (rather than anemometer) to detect pressure drop across the supply filter, rather than in just one point across the down flow. The pressure sensor must be encased in order to protect the sensor from temperature, humidity and other environmental phenomena that can impact the sensor's performance
The microprocessor must display the inflow and down flow air velocities in real-time on an LED display to ensure the user knows whether or not the cabinet is working under safe operating conditions.
The front window must be a 10" sash opening and be made of laminated safety glass to ensure containment of potentially hazardous samples in the case of accidental glass breakage.
All interior and exterior parts must be painted or smooth to ensure no risk of cuts to users or maintenance personnel.
The front of the cabinet must be angled 10° to help minimize glare on the window to the user, and to ensure that the user's posture is comfortable during a working session. Inadequate user ergonomics in a safety cabinet may lead to excessive fatigue, unsafe working habits and harmful consequences to user safety or product contamination.
The cabinet noise level must be less than 63 dB(A) for a 4 foot cabinet as measured in a sound proof room 12 inches in front of the cabinet and 15 inches above the work surface. Lower noise levels promote more comfortable and safer working habits of the user.
The Biosafety Cabinet should have microprocessor controller and same must be located on a slanted front panel so it is easy to see and reach from a seated working position in front of the cabinet.
The interior of the front window must be accessible for cleaning without requiring the user remove or support the window.
The biological safety cabinet must be capable of achieving current state-of-the-art in energy efficiency. A biological safety cabinet with lights on and fan at operating speed should consume less than 200 watts for a nominal four foot width and have a reduced energy mode for non-operational maintenance on containment in the work area.
The cabinet must automatically reduce fan/blower motor speed to 30% when the front window sash is in closed position to ensure reduced energy consumption when the cabinet is not is use.
In order to provide maximum effectiveness, efficiency and safety to laboratory Personnel, UV light must be programmable to allow for specific exposure times from 0 to 24 hours. The automatic shut off feature on the UV light saves money on replacement of the bulbs.
The Cabinet should have provision to fit taps for Vacuum, Water and Non Combustible Gas. Taps should be quoted as optional items
The Bio safety Cabinet should be NSF certified with listing on NSF website.
The Bio safety cabinet should incorporate HEPA filter of the class H 14 EN 1822 or better and having minimum efficiency of 99.995% at 0.3 µm particle size.



Dimension
Exterior 1500 H x 1300 W x 800 D; Interior 800 H x 1200 W x 500 D
Ventilation System Exhaust and Inflow air volume approx 300-350 CFM
Heat Emissions at 25°C should be approx 0.2 KW or lesser.
The cabinet Should be provided with Microprocessor controller and large LED display for inflow and Down flow air velocity and hours of operation, Audible and visual Alarms for HEPA filter failure, blower failure, airflow speed failure, Incorrect window position.
The BSC must incorporate an LED Indicator to indicate filter loading and should provide visual and audible alarm to indicate excessive HEPA filters loading which can result in unsafe airflows deviation from the NSF recommended inflow and down flows air velocity values measured in meters per second or foot per minute.
The cabinet should be provided with fixed / adjustable Height Stand, UV Light and one set of detachable arms rest, and one/two electrical outlet.
The Drain Pan of the BSC should be made of Stainless Steel. The drain pan should not be painted or coated.
The Bio safety cabinet should have dual side wall with negatively pressurized interstitial space.
The supplied instruments must have at least 3 years of warranty. The installation and training must be given by a Bangalore based Engineer, who would be available to rectify any issue within 2 working days of the notice. Details and contact number of this engineer must be given. Offered instrument's quality and userfriendly features will be confirmed from our previous experience or users from different institutes. Conclusion will be derived by internal technical team.

**Terms and Conditions:**

1. The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.
2. **first sealed cover – Cover I**, and super scribed as “**Techno-commercial Bid**” and should contain Complete Technical details of the Instrument offered (Specifications, Technical Parameters, Advantages, etc.,)
3. The **second sealed cover – Cover II** super scribed '**Price Bid**' should contain **only rates** (should be duly signed with seal and filled with date wherever necessary)
4. THESE TWO COVERS SHALL BE AGAIN PUT INTO A SINGLE WAX SEALED COVER super scribed “**Biosafety Cabinet.**” and should reach **INSTEM on or before 11-02-2021 before 5.30 P.M**”. This should be addressed to the Purchase Officer, inStem, NCBS campus, GKVK Post, Bellary Road, Bangalore – 65.
5. The tender to be quoted in foreign currencies & any other currencies approved/traded by RBI-USD/Euro/JPY/GBP/SGD/CAD/INR.
6. If the items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.
7. **Original Equipment Manufacturer or Authorised Dealers should provide Authorized Distribution Certificate along with the Quotation.**
8. **Kindly mention GEM Supplier ID on the Quotation.**
9. **Payment Term: 97% against supply; balance 3% on Installation subject to receipt of PBG for 3% Order Value.**





**INFORMATION TO TENDERERS**

**The Tender shall be evaluated under 2 (Two) Bid System**

**I Technical Bid**

**II Financial Bid**

**TECHNICAL SPECIFICATIONS & EVALUATION CRITERIA WITH MARKS FOR 2**

**PART TENDER FOR “Biosafety Cabinet -Qty 01 No.**

<b>Tender Specifications</b>	<b>Score</b>
The Bio safety cabinet should be of high quality, imported, Type A2 in which 70% Air should be re-circulated and 30% of the air should be exhausted	25
The Bio Safety Cabinet must include two DC motors. High power consuming /high heat output AC motors must not be used	
The motor must automatically adjust the airflow speed without the use of a damper to ensure continuous safe working conditions, even without maintenance adjustment , inflow and outflow should be independent control	7
In order to preserve safety to the user and the environment, the exhaust blower on the cabinet must continue operating when the supply blower stops working. If the exhaust blower should fail, the supply filter should also be turned off.	
In order to ensure consistent and reliable down flow velocity across the supply HEPA filter over the life of the cabinet, the cabinet must use a pressure sensor (rather than anemometer) to detect pressure drop across the supply filter, rather than in just one point across the down flow. The pressure sensor must be encased in order to protect the sensor from temperature, humidity and other environmental phenomena that can impact the sensor’s performance	10
The microprocessor must display the inflow and down flow air velocities in real-time on an LED display to ensure the user knows whether or not the cabinet is working under safe operating conditions.	
The front window must be a 10” sash opening and be made of laminated safety glass to ensure containment of potentially hazardous samples in the case of accidental glass breakage.	10
All interior and exterior parts must be painted or smooth to ensure no risk of cuts to users or maintenance personnel.	
The front of the cabinet must be angled 10° to help minimize glare on the window to the user, and to ensure that the user’s posture is comfortable during a working session. Inadequate user ergonomics in a safety cabinet may lead to excessive fatigue, unsafe working habits and harmful consequences to user safety or product contamination.	
The cabinet noise level must be less than 63 dB(A) for a 4 foot cabinet as measured in a sound proof room 12 inches in front of the cabinet and 15 inches above the work surface. Lower noise levels promote more comfortable and safer working habits of the user.	10
The Biosafety Cabinet should have microprocessor controller and same must be located on a slanted front panel so it is easy to see and reach from a seated working position in front of the cabinet.	
The interior of the front window must be accessible for cleaning without requiring the user remove or support the window.	
The biological safety cabinet must be capable of achieving current state-of-the-art in energy efficiency. A biological safety cabinet with lights on and fan at operating speed should consume less than 200 watts for a nominal four foot width and have a reduced energy mode for non-operational maintenance on containment in the work area.	10
The cabinet must automatically reduce fan/blower motor speed to 30% when the front window sash is in closed position to ensure reduced energy consumption when the cabinet is not is use.	
In order to provide maximum effectiveness, efficiency and safety to laboratory Personnel, UV light must be programmable to allow for specific exposure times from 0 to 24 hours. The automatic shut off feature on the UV light saves money on replacement of the bulbs.	



The Cabinet should have provision to fit taps for Vacuum, Water and Non Combustible Gas. Taps should be quoted as optional items	10
The Bio safety Cabinet should be NSF certified with listing on NSF website.	
The Bio safety cabinet should incorporate HEPA filter of the class H 14 EN 1822 or better and having minimum efficiency of 99.995% at 0.3 µm particle size.	
Dimension	
Exterior 1500 H x 1300 W x 800 D; Interior 800 H x 1200 W x 500 D	
Ventilation System Exhaust and Inflow air volume approx 300-350 CFM	
Heat Emissions at 25°C should be approx 0.2 KW or lesser.	5
The cabinet Should be provided with Microprocessor controller and large LED display for inflow and Down flow air velocity and hours of operation, Audible and visual Alarms for HEPA filter failure, blower failure, airflow speed failure, Incorrect window position.	
The BSC must incorporate an LED Indicator to indicate filter loading and should provide visual and audible alarm to indicate excessive HEPA filters loading which can result in unsafe airflows deviation from the NSF recommended inflow and down flows air velocity values measured in meters per second or foot per minute.	
The cabinet should be provided with fixed / adjustable Height Stand, UV Light and one set of detachable arms rest, and one/two electrical outlet.	13
The Drain Pan of the BSC should be made of Stainless Steel. The drain pan should not be painted or coated.	
The Bio safety cabinet should have dual side wall with negatively pressurized interstitial space.	
The supplied instruments must have at least 3 years of warranty. The installation and training must be given by a Bangalore based Engineer, who would be available to rectify any issue within 2 working days of the notice. Details and contact number of this engineer must be given. Offered instrument's quality and userfriendly features will be confirmed from our previous experience or users from different institutes. Conclusion will be derived by internal technical team.	
<b>Total</b>	<b>100</b>
<b>Evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.</b>	



**Bid Security Declaration**  
**(to be submitted on Company's Letter Head)**

I/We the undersigned hereby declare that if we withdraw or modify the bids during the period of its validity, or if we are awarded the contract and fail to sign the contract, or to submit a performance security before the deadline defined in the request for bids document or fail to execute the contract, we will be suspended for the period of time specified in the request for bids document from being eligible to submit bids for contracts with the entity that invited the bids.

Name and Signature  
of Authorized Signatory  
and Company Seal

Certificate for Local Content

“We (name of manufacturer) hereby confirm in respect of quoted item(s) that Local Content is equal to or more than 50% and come under “Class-I Local Supplier” Category. As being “Class – I Local Supplier”, we are eligible for Purchase Preference under “Make in India” Policy vide GOI Order No.P-45021/2/2017-PP (B.E.-II) dated 15.06.2017 (subsequently revised vide orders dated 28.05.2018, 29.05.2019 and 04.06.2020).

OR

“We (name of manufacturer) hereby confirm in respect of quoted item(s) that Local Content is more than 20% but less than 50% and come under “Class-II Local Supplier” Category.

The details of the location (s) at which the local value addition made is/are as under:

- 1.
- 2.
- 3.

\*Strike out whichever is not applicable

Date:

Seal & Signature of the Bidder