

Ref: INS/L-5623/2018-2019(Y)

Date: 18/09/2018

## ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following:

Sl. No	Cat. No	Item Description	Make/Model	Item Qty	UOM
1		-80 Degree Freezer (As per Specifications attached)		2.00	No.

**Remarks:** The Price quoted against this RFQ should be extended to Bangalore Life Science Cluster (BLiSc) of NCBS, InStem and C-CAMP for placing repeat order as per norms, by any one or all institutes of Bangalore Life Science Cluster (BLiSc)

**NOTE:** \*\*inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification. \*\*

The Tenders to be quoted in foreign currencies & any other currencies approved/traded by RBI - USD/Euro/JPY/GBP/SGD/CAD/INR.

1. The bids shall be enclosed in an envelope, and due date sealed duly marked "Tender for \_\_\_\_\_"  
Ref No: \_\_\_\_\_. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission Date and Time and inStem will not be responsible for the delay.

### 2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS

04/10/2018 till 4.30pm

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & Warranty period.


7. Delivery to be made to our Stores. Please indicate charges, if any extra. Transit Insurance should be done up to inStem Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/Specifications.

9. Payment: within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is super scribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs Duty (Except advalorem duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipment, technical instruments, equipment (including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.



**Institute for Stem Cell Biology and Regenerative Medicine**  
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5623/2018-2019(Y)

Date: 18/09/2018

12.If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with Xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act,1996 as amended from time to time shall apply. The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Institute for Stem Cell  
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to be 'Yesu R'.

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in





Ref No: INS/L-5623/2018-2019(Y)

### **Specification for -80 Freezer**

1. Freezer should have 700- 750 Lit capacity.
2. Temperature range should be -50°C to -85°C (1°C increment).
3. Freezer should have two independent compressor (Dual cooling) to provide a reliable -80°C ultra-low temperature environment.
4. Freezer should have various alarms including high/low temperatures, door ajar, power failure alarm and part replacement notification help to keep samples safe even in an emergency.
5. Supplier should supply complete stainless steel racks for entire freezer.
6. Freezer should have a thin Vacuum insulation panel which decreases the cabinet wall thickness by half and achieves up to 30% more storage capacity than conventionally insulated freezers.
7. The Freezer should have a clear large LCD Color Touch Display with Data log function
8. Should be a Filter less construction to eliminate the filter cleaning and maintenance.
9. Supplier should provide 3 years' warranty for freezer.

