Institute For Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5889/2018-2019(Y) Date: 04/12/2018

ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following:

SI.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		CO2 Incubators (Specifications Attached)		1.00	No.

Remarks: The Price quoted against this RFQ should be extended to Bangalore Life Science Cluster (BLiSc) of NCBS, Instem and C-CAMP for placing repeat order as per norms, by any one or all institutes of Bangalore Life Science Cluster (BLiSc) **inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification. **

Note: The Tenders to be quoted in foreign currencies & any other currencies approved/traded by RBI - USD/Euro/JPY/GBP/SGD/CAD/INR.

1.	The	bids	shall	be	enclosed	in	an	envelope	,	and	due	date	sealed	duly	marked	"Tender	for
" Ref No :												Th	e bids s	hould	be addres	ssed and to	o be
ma	iled to	TT"	E HE.	AD-l	PURCHA	SE"	. Th	e bids are l	iat	ole to	be rej	ected	if the sea	aled er	velope is	not addre	ssed
to	"THE	HEA	AD-PU	RCI	HASE" w	ith 7	end	er Ref No	an	d Iten	n Des	criptio	n and d	ue date	e. The bid	ls delivere	ed in
pe	rson sl	nall be	e drop	ped	in Purchas	e S	ectio	n. If the b	ids	are s	sent tl	rough	ourie	or m	ail, it sh	ould reacl	h by
sul	omissio	on Da	te and	Time	e and inSte	em v	vill r	not be respo	ons	ible f	or the	delay					

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 18/12/2018 till 5.30 p.m.

- 3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.
- 4. The Validity of your quotation should be for 60 days from the date.
- 5.All duties,taxes,surcharge and cess as currently applicable must be sated in your quotation,seperately.Otherwise your quote is liable to be rejected.
- 6. Your quotation should indicate delivery period & Warranty period.
- 7.Delivery to be made to our Stores.Please indicate charges, if any extra.Transit Insurance should be done upto inStem Stores.
- 8.If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, Please state so and then offer alternative to quality/Specifications.
- 9.Payment :within one month after delivery & acceptance/satisfactory installation.
- 10.Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.
- 11. Since we are a public funded research institution, we are exempted from paying Customs Duty(Except advolerum duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.

If the items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.

Institute For Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5889/2018-2019(Y)

Date: 04/12/2018

12.If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Concillation Act, 1996 as amended from time to time shall apply. The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully
For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No.:

91-80-23666343/344/345/346

Email Id:

purchase@ncbs.res.in

Fax: 91-80-23636662

Website: www.ncbs.res.in



Ref No: INS/L-5889/2018-2019(Y)

Specifications for Air jacketed CO₂ incubator

- 1. The system should be Air jacketed C0₂ incubator.
- 2. The system should have water tray inside to maintain the moisture
- 3. The system should be microprocessor controlled C0₂ incubator It should have touch screen LCD controller with user-friendly control system and GUI.
- 4. The system should have the capacity of 170 Ltr(\pm 5%)
- 5. The system should have two doors and inner door should be glass door to seal the chamber with Silicon gasket.
- 6. The system should have the seamless stainless steel (SS) inner chamber.
- 7. The system should be supplied with minimum 3 nos of adjustable SS shelves.
- 8. The system should have the Infrared $C0_2$ sensor with the $C0_2$ range 1-20% with accuracy 0.1%.
- 9. The system should have the Operating temperature range from ambient to 50°C. The fluctuation/variation at 37°C should not be more than 0.1°C.
- 10. The temperature accuracy and uniformity should be less than 0.1 °C
- 11. The system should have the tracking alarm $\pm 1^{\circ}$ C deviation from set point.
- 12. Sensor must be able to withstand high temperatures during direct heat sterilization
- 13. The system should have Built in on-demand high temperature decontamination cycle (140°C and above)
- 14. The system should have the option to mute the audible alarms
- 15. The system should be suitable for single phase 230 V, 50 Hz Power supply.
- 16. The system should have onboard Diagnostic interface to show system parameters and functions.
- 17. The system should be equipped with buzzer and visual alarm for power failure, C0₂, and temperature deviation against the setup temperature.
- 18. The incubator shaker should be quoted with a suitable stackable kit with the main offer.
- 19. The system should operate at 230V, 50 Hz AC power.
- 20. The system should be supplied with all the accessories required to function.
- 21. The system should be offered with minimum 3-year warranty.
- 22. The System should be CE/ISO Approved.
- 23. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required).

