



Ref: INS/L-5905/2017-2018(Y)

Date : 05/12/2017

ENQUIRY

Dear Sirs,
Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Small Shaker incubator (Specifications attached)	any	2.00	No.

Remarks :The Tenders can also to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Note :

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for _____" Ref No : _____. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.
2. **DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 25/12/2017 till 5.30 p.m.**
3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.
4. The Validity of your quotation should be for 60 days from the date.
- 5.All duties,taxes,surcharge and cess as currently applicable must be sated in your quotation,seperately.Otherwise your quote is liable to be rejected.
- 6.Your quotation should indicate delivery period & Warranty period.
- 7.Delivery to be made to our Stores.Please indicate charges,if any extra.Transit Insurance should be done upto inStem Stores.
- 8.If you are unable to supply the quality,specifications or brand as mentioned in our enquiry,Please state so and then offer alternative to quality/Specifications.
- 9.Payment :within one month after delivery & acceptance/satisfactory installation.
- 10.Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.
- 11.Since we are a public funded research institution,we are exempted from paying Customs Duty(Except advolerum duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments,technical instruments,equipments(including computers),their accessories,spares,consumables and software.Hence,please offer your prices taking this option into consideration.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5905/2017-2018(Y)

Date : 05/12/2017

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to read 'Jesu R'.

JESU R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



Ref No: INS/L-5905/2017-2018(Y)

Specifications for Small shaker incubator

1. Shaking Frequency : 30-350 RPM
 2. Shaking Deviation : ± 2 RPM
 3. Shaking Display : Digital LED
 4. Shaking Change via : Soft Touch Keypad
 5. Shaking Motion : Orbital
 6. Diameter of Orbital Motion : 25 mm
 7. Platform Dimension (W x D) : 420 x 420 mm
 8. Movement of the shaker comprises Triple Eccentric Counter Balanced drive with 7 permanently Lubricated Ball Bearings
 9. Drive : BLAC
 10. Universal Aluminum Platform can accommodate standard clamps of sizes from 25 ml upto 2 L
 11. Temperature Range : Ambient 7°C to 70°C
 12. Temperature Accuracy : $\pm 0.1^\circ$ C
 13. Permissible Ambient Temperature : 5°C to 40°C
 14. Temperature Sensor : Pt100
 15. Temperature Controller : PID
 16. Temperature Display : Digital LED
 17. Temperature Change via : Soft Touch Keypad
 18. Front open door with large viewing mirror
 19. Material of Construction : Interior and Exterior MS Powder coated
 20. Internal Dimensions (W x D x H) : (490x 500x 412) mm
 21. External Dimensions (W x D x H) : (525 x 730x 800) mm
 22. Can be stacked upto 2 Levels using Stack Rack / Stacking Kit
 23. Audible & Visual alarm for temperature deviation.
 24. Thermostat cut-off for run away safety
 25. Provided with RCCB to prevent from electrical hazard
 26. Non Volatile Memory with automatic Power failure restart
 27. Unit has facility of proximity sensor to control the shaker on/off position
 28. Electrical Requirement : 230V, 6 Amps, 50 Hz, 1 Φ
- CE certified to comply with the international safety measures

