

Ref: INS/L-5937/2018-2019(Y)

Date: 12/12/2018

## ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following:

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Magnetic Stirrer with Hot Plate (Specifications Attached)		2.00	Nos.

**Remarks :** The Price quoted against this RFQ should be extended to Bangalore Life Science Cluster (BLiSc) of NCBS, InStem and C-CAMP for placing repeat order as per norms, by any one or all institutes of Bangalore Life Science Cluster (BLiSc). \*\*inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification. \*\*

**Note : The Tenders to be quoted in foreign currencies & any other currencies approved/traded by RBI - USD/Euro/JPY/GBP/SGD/CAD/INR.**

1. The bids shall be enclosed in an envelope, and due date sealed duly marked "Tender for \_\_\_\_\_" Ref No : \_\_\_\_\_. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission Date and Time and inStem will not be responsible for the delay.

### 2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS

26/12/2018 till 5.30 p.m.

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & Warranty period.

7. Delivery to be made to our Stores. Please indicate charges, if any extra. Transit Insurance should be done upto inStem Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment : within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs Duty (Except ad valorem duty of 5% + 2% cess and CVD @ 4% vide Notification No. 51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.



Ref: INS/L-5937/2018-2019(Y)

Date : 12/12/2018

**12.If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with Xerox copy of the DGS&D order.**

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Cancellation Act,1996 as amended from time to time shall apply. The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Institute For Stem Cell  
Biology and Regenerative Medicine



Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666341/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in



INS/L-5937/2018-2019(Y)

### **Magnetic Stirrer hot plate**

1. The system should have both stirrer and heater and should have separate control.
2. The should have hotplate and should be ceramic coated.
3. The system should have the speed range of 60 – 1200 RPM with accuracy of  $\pm 2$  rpm.
4. The system should have the Maximum temperature 300°C with minimum 750W heater.
5. The system should have microcontroller based temperature control Option for temperature accuracy of  $\pm 1^\circ\text{C}$ .
6. The system should be able to accommodate maximum 2/3 liters' flask and should have stirring capacity of 10Lt water.
7. The system should have an independent safety circuit with separate circuit switches to cut off the heating if the normal hotplate temperature is exceeded by 20°C
8. The system should be suitable for Single phase 230v/50hz Power supply.
9. The system should be CE/ISO Certified
10. The system should have minimum 3-year warranty.
11. The system should be supplied with all the accessories required to function

