

Ref: INS/L-6124/2017-2018(N)

Date : 22/01/2018

### ENQUIRY

Dear Sirs,  
Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Vacuum Drying oven (As per the specification attached)	Any	1.00	Nos.

**Remarks: The Tenders can also be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.**

**NOTE: \*\*inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification.\*\***

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for \_\_\_\_\_" Ref No : \_\_\_\_\_. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS  
07/02/2018 till 5.30pm**

**3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.**

4. The Validity of your quotation should be for 60 days from the date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & Warranty period.

7. Delivery to be made to our Stores. Please indicate charges, if any extra. Transit Insurance should be done upto inStem Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, Please state so and then offer alternative to quality/Specifications.

9. Payment : within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs Duty (Except ad valorem duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.



**Institute For Stem Cell Biology and Regenerative Medicine**  
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-6124/2017-2018(N)

Date : 22/01/2018

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell  
Biology and Regenerative Medicine

*Namratha A*  
22/01/18

Namratha A

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



# INSTITUTE FOR STEM CELL BIOLOGY AND REGENERATIVE MEDICINE

Ref No: INS/F-6124/2017-2018(N)

## Specification for vacuum Drying oven

1. The system should be microprocessor controlled Incubator It should have LED Display with easy icons and user-friendly control system.
2. The system should have the capacity of 70 liters  $\pm$  5%
3. The system should have temperature range from ambient 5°C to 250°C.
4. The system should have the preset function for atleast 2 commonly used temperature settings.
5. The fluctuation/variation at 37°C should not be more than 0.5°C
6. The system should have facility to calibrate the temperature via keypad
7. The system should be supplied with 2 nos of adjustable shelves
8. The system should have the timer range 1 min to 99 hr 59 min,
9. The system should have the Air circulation by a fan to maintain temperature uniformity.
10. The system should have over temperature and over current protection circuit.
11. The system should have the Corrosion resistance stainless steel interior and epoxy-powder coated exterior.
12. The system should have the automatic restart after the power interruption
13. The system should have silicon rubber door seal gasget to ensure a high vacuum inside the chamber
14. The system should have tempered glass for easy monitoring of samples
15. Electronic and mechanical components should be enclosed, protected from accidental spills.
16. The system should be quoted with the suitable vacuum pump with the main offer.
17. The system should have the vacuum range of 0 to 0.1MPa with analog gauge
18. The related accessories and consumables should be quoted with the main offer.
19. The system should be equipped with RS-232 interface for external data collection.
20. The system should operate at 230V, 50 Hz AC power.
21. The system should be supplied with all the accessories required to function.
22. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required).
23. The system should be supplied with 3 years warranty.
24. The System should be CE/ISO Approved

