

Ref: INS/L-5967/2018-2019(Y)

Date : 03/01/2019

ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following:

SL No	Item Description	Make/Model	Item Qty	UOM
1	Microscope (Specifications attached)		1.00	No.

Remarks: The Price quoted against this RFQ should be extended to Bangalore Life Science Cluster (BLiSc) of NCBS, InStem and C-CAMP for placing repeat order as per norms, by any one or all institutes of Bangalore Life Science Cluster (BLiSc)

NOTE: **inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% For items supplied for research purpose. The offer should be submitted after fully considering the above notification. **

The Tenders to be quoted in foreign currencies & any other currencies approved/traded by RBI - USD/Euro/JPY/GBP/SGD/CAD/INR.

1. The bids shall be enclosed in an envelope, and due date sealed duly marked "Tender for _____" Ref No : _____. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission Date and Time and inStem will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS

10/01/2019 till 5.30 p.m.

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & Warranty period.

7. Delivery to be made to our Stores. Please indicate charges, if any extra. Transit Insurance should be done up to inStem Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/Specifications.

9. Payment : within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscripted on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs Duty (Except advalorem duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments(including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.

If the items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penal-ty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.



Institute For Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5967/2018-2019(Y)

Date :03/01/2019

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with Xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Institute for Stem Cell
Biology and Regenerative Medicine

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



INSTITUTE FOR STEM CELL BIOLOGY AND REGENERATIVE MEDICINE

Ref: INS/L-5967/2018-2019(Y)

SPECIFICATIONS for Microscope (Fluorescent imaging)		
Main body	Optical system	Infinity-corrected CFI60
	Output port	4 Manual positions Eyepiece 100%, left 100%, right 100%, option (to eyepiece 20%/left 80% or eyepiece 20%/right 80%) Can add ports by use of back port unit.
	Focusing unit	Manual drive, Coarse/fine focusing knob, 10mm stroke
	Stage up	Available
Tube body	Binocular tube	Binocular S tube (field number 22),
Transmitted illumination	Pillar for transmitted illumination	Condenser vertical stroke: 66mm, Backward tilting up to 25 degrees, With field diagram and refocusing mechanism 2 filter slot positions (4 filter position option is also available with Filter Slider for transmitted illumination). High power LED or 100W halogen bulb (pre-centered)
Condenser	Condenser turret (TC-C-TC)	7 manual positions (ø37mm x4, ø39mm x3), LWD/ELWD/CLWD/NAMC condenser lenses are supported
	Condenser lens	LWD (W.D.=30mm, NA=0.52), ELWD (W.D.=75mm, NA=0.3).
Stage	Stage	Stroke X: ±57mm, Stroke Y: ±36.5mm, Adjustable stroke range (3 levels) with adjusting pin, Long/middle/short handle options available
Objectives	Objectives	Achromat 4x/0.1/30mm, Achromat ADL 10x/0.25/5.2mm/Ph1, Achromat ADL LWD 20x/0.4/3.0mm/Ph1, Fluor/Fluorite ELWD ADM 40x/0.60/3.6-2.8mm/Ph2, Plan Apochromat/APO 60X/0.95/W.D. 0.15 mm.
Nosepiece	Sextuple nosepiece	6 manual positions, Simple waterproof structure
Epi-filter turret	Intelligent epi filter turret	6 manual positions, Manual shutter
Epi-fluorescence attachment	EPI-FL module	Supports fiber illuminator; includes 2-position filter slider and aperture diaphragm
	Field stop slider	Circular, rectangular, square aperture options
Epi-fluorescence light source	Illumination	130W high vapour mercury lamp with 2000 working hours or equivalent LED illumination for Blue, Green & Red excitations
Photomicrography Attachment	Camera	Digital CCD/CMOS image sensor, 2880x2048 pixel 5.9 megapixels, 1/1.8 inch (6.912x4.915mm), "All pixels" mode : 2880x2048 pixel, "2 Vertical and 2 horizontal pixels average" mode : 1440x1024 pixel, ROI 1/2 mode : 1440x1024 pixel, "All pixels" mode (2880x2048 pixel) : 15fps, "2 Vertical and 2 horizontal pixels average" mode (1440x1024 pixel) : 30fps, 100µsec to 30 sec, USB3.0, Micro-B Receptacle.
Software	Software	Acquisition and device control up to three-dimensional acquisition, Image Acquisition, Measurements, Time Lapse Imaging, Z-Stack, Annotation, Large Image, Macro, Report Generator facility, Vector layer & Multi-Dimensional File Format
Suitable PC for capturing & storing image should be provided along with the Microscope.		

MICROSCOPE, CAMERA & SOFTWARE SHOULD BE FROM SAME MANUFACTURER FOR BETTER COMPATIBILITY.

